

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF010522\
 Data File : BF126490.D
 Acq On : 5 Jan 2022 11:42
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 06 01:07:39 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 05 01:57:07 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	82	0.00
2	1,4-Dioxane	40.000	38.380	4.0	75	0.00
3	Pyridine	40.000	37.470	6.3	74	-0.01
4	n-Nitrosodimethylamine	40.000	33.846	15.4	68	-0.01
5 S	2-Fluorophenol	80.000	77.671	2.9	79	0.00
6	Aniline	40.000	36.867	7.8	74	0.00
7 S	Phenol-d6	80.000	76.000	5.0	78	0.00
8	2-Chlorophenol	40.000	39.914	0.2	81	0.00
9	Benzaldehyde	40.000	32.775	18.1	71	0.00
10 C	Phenol	40.000	37.213	7.0	75	0.00
11	bis(2-Chloroethyl)ether	40.000	37.284	6.8	76	0.00
12	1,3-Dichlorobenzene	40.000	40.629	-1.6	83	0.00
13 C	1,4-Dichlorobenzene	40.000	40.660	-1.6	82	0.00
14	1,2-Dichlorobenzene	40.000	38.461	3.8	83	0.00
15	Benzyl Alcohol	40.000	36.901	7.7	74	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.917	7.7	75	0.00
17	2-Methylphenol	40.000	38.145	4.6	78	0.00
18	Hexachloroethane	40.000	39.211	2.0	78	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	34.604	13.5	74	0.00
20	3+4-Methylphenols	40.000	35.736	10.7	78	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	81	0.00
22	Acetophenone	40.000	37.948	5.1	78	0.00
23 S	Nitrobenzene-d5	80.000	74.979	6.3	74	0.00
24	Nitrobenzene	40.000	37.380	6.5	73	0.00
25	Isophorone	40.000	36.421	8.9	74	0.00
26 C	2-Nitrophenol	40.000	41.374	-3.4	80	0.00
27	2,4-Dimethylphenol	40.000	40.508	-1.3	79	0.00
28	bis(2-Chloroethoxy)methane	40.000	37.655	5.9	75	0.00
29 C	2,4-Dichlorophenol	40.000	41.579	-3.9	84	0.00
30	1,2,4-Trichlorobenzene	40.000	41.407	-3.5	83	0.00
31	Naphthalene	40.000	40.091	-0.2	81	0.00
32	Benzoic acid	40.000	42.312	-5.8	94	0.00
33	4-Chloroaniline	40.000	39.825	0.4	79	0.00
34 C	Hexachlorobutadiene	40.000	41.274	-3.2	83	0.00
35	Caprolactam	40.000	38.046	4.9	79	0.00
36 C	4-Chloro-3-methylphenol	40.000	37.969	5.1	76	0.00
37	2-Methylnaphthalene	40.000	40.476	-1.2	81	0.00
38	1-Methylnaphthalene	40.000	40.464	-1.2	82	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	81	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	43.257	-8.1	86	0.00
41 P	Hexachlorocyclopentadiene	40.000	43.075	-7.7	77	0.00
42 S	2,4,6-Tribromophenol	80.000	83.841	-4.8	81	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.534	-6.3	82	0.00
44	2,4,5-Trichlorophenol	40.000	42.060	-5.2	83	0.00
45 S	2-Fluorobiphenyl	80.000	77.251	3.4	84	0.00
46	1,1'-Biphenyl	40.000	41.572	-3.9	82	0.00
47	2-Chloronaphthalene	40.000	41.350	-3.4	81	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF010522\
 Data File : BF126490.D
 Acq On : 5 Jan 2022 11:42
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 06 01:07:39 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 05 01:57:07 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	35.753	10.6	69	0.00
49	Acenaphthylene	40.000	41.033	-2.6	81	0.00
50	Dimethylphthalate	40.000	39.622	0.9	79	0.00
51	2,6-Dinitrotoluene	40.000	39.502	1.2	77	0.00
52 C	Acenaphthene	40.000	40.657	-1.6	79	0.00
53	3-Nitroaniline	40.000	37.899	5.3	74	0.00
54 P	2,4-Dinitrophenol	40.000	34.672	13.3	69	0.00
55	Dibenzofuran	40.000	40.211	-0.5	79	0.00
56 P	4-Nitrophenol	40.000	37.271	6.8	69	0.00
57	2,4-Dinitrotoluene	40.000	39.074	2.3	74	0.00
58	Fluorene	40.000	39.353	1.6	80	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.235	-3.1	80	0.00
60	Diethylphthalate	40.000	37.802	5.5	74	0.00
61	4-Chlorophenyl-phenylether	40.000	40.048	-0.1	81	0.00
62	4-Nitroaniline	40.000	35.730	10.7	70	0.00
63	Azobenzene	40.000	35.411	11.5	69	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	72	0.00
65	4,6-Dinitro-2-methylphenol	40.000	45.156	-12.9	73	0.00
66 c	n-Nitrosodiphenylamine	40.000	43.623	-9.1	79	0.00
67	4-Bromophenyl-phenylether	40.000	45.150	-12.9	80	0.00
68	Hexachlorobenzene	40.000	43.372	-8.4	78	0.00
69	Atrazine	40.000	42.295	-5.7	71	0.00
70 C	Pentachlorophenol	40.000	41.895	-4.7	69	0.00
71	Phenanthrene	40.000	40.984	-2.5	74	0.00
72	Anthracene	40.000	40.859	-2.1	74	0.00
73	Carbazole	40.000	38.836	2.9	71	0.00
74	Di-n-butylphthalate	40.000	36.556	8.6	65	0.00
75 C	Fluoranthene	40.000	36.488	8.8	66	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	66	0.00
77	Benzydine	40.000	35.672	10.8	57	0.00
78	Pyrene	40.000	40.326	-0.8	64	0.00
79 S	Terphenyl-d14	80.000	79.234	1.0	65	0.00
80	Butylbenzylphthalate	40.000	35.811	10.5	56	0.00
81	Benzo(a)anthracene	40.000	40.466	-1.2	65	0.00
82	3,3'-Dichlorobenzidine	40.000	44.234	-10.6	73	0.00
83	Chrysene	40.000	42.093	-5.2	68	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.060	-0.2	65	0.00
85 c	Di-n-octyl phthalate	40.000	48.141	-20.4#	75	0.00
86 I	Perylene-d12	20.000	20.000	0.0	95	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	43.713	-9.3	98	0.00
88	Benzo(b)fluoranthene	40.000	36.455	8.9	88	0.00
89	Benzo(k)fluoranthene	40.000	38.640	3.4	85	0.00
90 C	Benzo(a)pyrene	40.000	40.686	-1.7	93	0.00
91	Dibenzo(a,h)anthracene	40.000	44.113	-10.3	99	0.00
92	Benzo(g,h,i)perylene	40.000	43.548	-8.9	97	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF010522\
Data File : BF126490.D
Acq On : 5 Jan 2022 11:42
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jan 06 01:07:39 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 05 01:57:07 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 1