



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF010622\
 Data File : BF126511.D
 Acq On : 6 Jan 2022 10:18
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122921.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Jan 05 01:57:07 2022






AutoFind: Scans 1611, 1612, 1613; Background Corrected with Scan 1606

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	42.2	32993	PASS
68	69	0.00	2	1.7	483	PASS
69	198	0.00	100	37.4	29235	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	52.8	41284	PASS
197	198	0.00	2	0.5	368	PASS
198	198	100	100	100.0	78211	PASS
199	198	5	9	6.9	5389	PASS
275	198	10	60	24.8	19396	PASS
365	198	1	100	3.6	2796	PASS
441	198	0.01	100	16.9	13237	PASS
442	442	50	100	100.0	84323	PASS
443	442	15	24	17.9	15128	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF010622\
Data File : BF126511.D
Acq On : 6 Jan 2022 10:18
Operator : JU/CG
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Jan 07 11:25:10 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 05 01:57:07 2022
Response via : Initial Calibration



TIC: BF126511.D\data.ms

(77) Benzidine

12.639min (+ 0.000) 0.00 ng




response 499302

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.80	14.00
183.00	10.20	10.37
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF010622\
Data File : BF126511.D
Acq On : 6 Jan 2022 10:18
Operator : JU/CG
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Jan 07 11:25:10 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 05 01:57:07 2022
Response via : Initial Calibration



TIC: BF126511.D\data.ms

(70) Pentachlorophenol (C)

11.104min (-0.012) 45254.12 ng

response 70851

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	65.40	67.79
264.00	63.00	64.15
0.00	0.00	0.00

Instrument :

BNA_F

ClientSampleId :

DFTPP

DDT Breakdown

Date	Instrument Name	DFTPP Data File
1/6/2022	BNA_F	<u>BF126511.D</u>
Compound Name	Response	Retention Time
DDT	186396	13.463
DDD	2315	13.163
DDE	241	12.827
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
2556	188952	1.35