


Data Path : Z:\HPCHEM1\BNA\_F\DATA\BF010815\  
 Data File : BF076774.D  
 Acq On : 8 Jan 2015 12:33  
 Operator : IZ / TP  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF121714.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Thu Jan 08 09:08:31 2015



AutoFind: Scans 1520, 1521, 1522; Background Corrected with Scan 1516

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 41.4      | 10437   | PASS             |
| 68          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 69          | 198          | 0.00         | 100          | 41.1      | 10338   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 127         | 198          | 10           | 80           | 52.9      | 13311   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 25182   | PASS             |
| 199         | 198          | 5            | 9            | 6.2       | 1562    | PASS             |
| 275         | 198          | 10           | 60           | 22.1      | 5571    | PASS             |
| 365         | 198          | 1            | 100          | 2.9       | 738     | PASS             |
| 441         | 442          | 0.01         | 24           | 15.9      | 3669    | PASS             |
| 442         | 198          | 50           | 100          | 91.5      | 23038   | PASS             |
| 443         | 442          | 15           | 24           | 22.1      | 5087    | PASS             |