

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF010917\
 Data File : BF092156.D
 Acq On : 9 Jan 2017 23:43
 Operator : UM/SJ
 Sample : I1045-01 2X
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 P3-SP1

Manual Integrations
 APPROVED

Sohil
 1/10/2017 10:01:41 AM

Quant Time: Jan 10 03:38:42 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF122116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 06 15:43:51 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.64	152	143855	20.00	ng	0.00
21) Naphthalene-d8	7.92	136	515794	20.00	ng	-0.01
38) Acenaphthene-d10	9.68	164	239498	20.00	ng	0.00
63) Phenanthrene-d10	11.16	188	405071	20.00	ng	0.00
75) Chrysene-d12	13.80	240	247544	20.00	ng	0.01
86) Perylene-d12	15.17	264	176748	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.25	112	222307	25.80	ng	0.01
7) Phenol-d6	6.32	99	409334	38.92	ng	0.01
23) Nitrobenzene-d5	7.20	82	230424	24.84	ng	-0.01
41) 2,4,6-Tribromophenol	10.47	330	25926	13.08	ng	0.00
44) 2-Fluorobiphenyl	9.01	172	419211	27.07	ng	0.00
78) Terphenyl-d14	12.74	244	327892	32.12	ng	0.00
Target Compounds						Qvalue
10) Phenol	6.33	94	25321	2.11	ng	78
49) Dimethylphthalate	9.41	163	62427	4.08	ng	99
70) Phenanthrene	11.18	178	120326	5.48	ng	96
77) Pyrene	12.60	202	91923	5.06	ng	97
80) Benzo(a)anthracene	13.79	228	29941	2.14	ng	# 78
82) Chrysene	13.82	228	32289m	2.30	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF010917\
Data File : BF092156.D
Acq On : 9 Jan 2017 23:43
Operator : UM/SJ
Sample : I1045-01 2X
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
P3-SP1

Manual Integrations
APPROVED

Sohil
1/10/2017 10:01:41 AM

Quant Time: Jan 10 03:38:42 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF122116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 06 15:43:51 2017
Response via : Initial Calibration

