

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF010917\
 Data File : BF092160.D
 Acq On : 10 Jan 2017 1:31
 Operator : UM/SJ
 Sample : I1045-05 2X
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 RAP-SP3

Manual Integrations
 APPROVED

Sohil
 1/10/2017 10:01:50 AM

Quant Time: Jan 10 03:53:56 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF122116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 06 15:43:51 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.64	152	136123	20.00	ng	0.00
21) Naphthalene-d8	7.93	136	505942m	20.00	ng	0.00
38) Acenaphthene-d10	9.68	164	288580	20.00	ng	0.00
63) Phenanthrene-d10	11.16	188	435969	20.00	ng	0.00
75) Chrysene-d12	13.80	240	204143	20.00	ng	0.01
86) Perylene-d12	15.18	264	153834	20.00	ng	0.03
System Monitoring Compounds						
5) 2-Fluorophenol	5.25	112	287534	35.27	ng	0.01
7) Phenol-d6	6.31	99	352090	35.37	ng	0.00
23) Nitrobenzene-d5	7.20	82	157034	17.26	ng	-0.01
41) 2,4,6-Tribromophenol	10.47	330	60609	25.38	ng	0.00
44) 2-Fluorobiphenyl	9.01	172	376091	16.27	ng	0.00
78) Terphenyl-d14	12.75	244	238585	28.34	ng	0.00
Target Compounds						
10) Phenol	6.33	94	22943	2.02	ng	Qvalue # 78
31) Naphthalene	7.94	128	53746	2.32	ng	99
37) 2-Methylnaphthalene	8.64	142	59286	2.83	ng	98
51) Acenaphthene	9.72	154	44176	2.71	ng	94
57) Fluorene	10.23	166	41056	2.56	ng	91
70) Phenanthrene	11.18	178	187064	7.91	ng	97
74) Fluoranthene	12.37	202	184531	5.54	ng	97
77) Pyrene	12.60	202	147325	9.84	ng	98
80) Benzo(a)anthracene	13.79	228	47215	4.10	ng	# 87
82) Chrysene	13.82	228	59760	5.17	ng	# 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF010917\
Data File : BF092160.D
Acq On : 10 Jan 2017 1:31
Operator : UM/SJ
Sample : I1045-05 2X
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
RAP-SP3

Manual Integrations
APPROVED

Sohil
1/10/2017 10:01:50 AM

Quant Time: Jan 10 03:53:56 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF122116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 06 15:43:51 2017
Response via : Initial Calibration

