

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF010917\
 Data File : BF092161.D
 Acq On : 10 Jan 2017 1:57
 Operator : UM/SJ
 Sample : I1045-06 2X
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 RAP-SP4

Manual Integrations
 APPROVED

Sohil
 1/10/2017 10:01:53 AM

Quant Time: Jan 10 04:00:59 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF122116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 06 15:43:51 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.64	152	155285	20.00	ng	0.00
21) Naphthalene-d8	7.93	136	591601	20.00	ng	0.00
38) Acenaphthene-d10	9.68	164	302947	20.00	ng	0.00
63) Phenanthrene-d10	11.16	188	437045	20.00	ng	0.00
75) Chrysene-d12	13.80	240	241230	20.00	ng	0.01
86) Perylene-d12	15.18	264	182393	20.00	ng	0.03
System Monitoring Compounds						
5) 2-Fluorophenol	5.25	112	337565	36.30	ng	0.01
7) Phenol-d6	6.31	99	433913	38.22	ng	0.00
23) Nitrobenzene-d5	7.20	82	178518	16.78	ng	-0.01
41) 2,4,6-Tribromophenol	10.47	330	63014	25.13	ng	0.00
44) 2-Fluorobiphenyl	9.01	172	422398	18.47	ng	0.00
78) Terphenyl-d14	12.75	244	260806	26.22	ng	0.00
Target Compounds						Qvalue
49) Dimethylphthalate	9.41	163	52548	2.71	ng	99
70) Phenanthrene	11.18	178	125371	5.29	ng	98
74) Fluoranthene	12.37	202	200703	6.37	ng	98
77) Pyrene	12.60	202	156198	8.83	ng	98
80) Benzo(a)anthracene	13.79	228	53288	3.91	ng	# 87
82) Chrysene	13.82	228	69260m	5.07	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF010917\
Data File : BF092161.D
Acq On : 10 Jan 2017 1:57
Operator : UM/SJ
Sample : I1045-06 2X
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
RAP-SP4

Manual Integrations
APPROVED

Sohil
1/10/2017 10:01:53 AM

Quant Time: Jan 10 04:00:59 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF122116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 06 15:43:51 2017
Response via : Initial Calibration

