


Data Path : Z:\HPCHEM1\BNA F\DATA\BF010917\
 Data File : BF092133.D
 Acq On : 9 Jan 2017 12:15
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF122116.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Jan 06 15:43:51 2017



AutoFind: Scans 834, 835, 836; Background Corrected with Scan 830

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.3	51951	PASS
68	69	0.00	2	1.5	786	PASS
69	198	0.00	100	34.6	53992	PASS
70	69	0.00	2	0.6	323	PASS
127	198	10	80	47.5	74157	PASS
197	198	0.00	2	0.6	969	PASS
198	198	100	100	100.0	156017	PASS
199	198	5	9	6.4	9980	PASS
275	198	10	60	24.4	38029	PASS
365	198	1	100	2.8	4322	PASS
441	442	0.01	24	15.7	20395	PASS
442	198	50	100	83.4	130050	PASS
443	442	15	24	18.8	24405	PASS