

Data Path : Z:\HPCHEM1\BNA F\DATA\BF010918\
 Data File : BF101969.D
 Acq On : 9 Jan 2018 22:51
 Operator : SJ/JU
 Sample : J1043-16
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-03-COMP

Manual Integrations
 APPROVED

Sohil
 1/10/2018 11:45:23 AM

Quant Time: Jan 10 00:14:52 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF010918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 09 15:20:36 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.73	152	212894	20.00	ng	0.00
21) Naphthalene-d8	8.01	136	761930	20.00	ng	0.00
38) Acenaphthene-d10	9.76	164	261742	20.00	ng	0.00
63) Phenanthrene-d10	11.25	188	424750	20.00	ng	0.00
75) Chrysene-d12	13.88	240	413198	20.00	ng	0.00
86) Perylene-d12	15.30	264	284622	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.34	112	1470654	97.55	ng	0.02
7) Phenol-d6	6.36	99	1665037	92.56	ng	0.01
23) Nitrobenzene-d5	7.29	82	983648	75.86	ng	0.00
41) 2,4,6-Tribromophenol	10.55	330	203095	88.91	ng	0.00
44) 2-Fluorobiphenyl	9.09	172	1112979	66.43	ng	0.00
78) Terphenyl-d14	12.83	244	842589	42.34	ng	0.00
Target Compounds						
10) Phenol	6.37	94	87198	4.289	ng	88
49) Dimethylphthalate	9.48	163	107492	5.119	ng	100
70) Phenanthrene	11.27	178	57388	2.313	ng	95
74) Fluoranthene	12.45	202	107083	4.387	ng	98
77) Pyrene	12.68	202	102254	2.800	ng	98
83) Bis(2-ethylhexyl)phthalate	13.87	149	49799	2.590	ng	# 99
87) Benzo(b)fluoranthene	14.89	252	48777m	2.628	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF010918\
Data File : BF101969.D
Acq On : 9 Jan 2018 22:51
Operator : SJ/JU
Sample : J1043-16
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-03-COMP

Manual Integrations
APPROVED

Sohil
1/10/2018 11:45:23 AM

Quant Time: Jan 10 00:14:52 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF010918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 09 15:20:36 2018
Response via : Initial Calibration

