

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF011017\
 Data File : BF092196.D
 Acq On : 10 Jan 2017 21:44
 Operator : UM/SJ
 Sample : I1098-03
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TP3-COMP13

Quant Time: Jan 10 23:36:12 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF122116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 06 15:43:51 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.62	152	145184	20.00	ng	-0.02
21) Naphthalene-d8	7.91	136	630791	20.00	ng	-0.02
38) Acenaphthene-d10	9.67	164	343425	20.00	ng	-0.01
63) Phenanthrene-d10	11.14	188	583909	20.00	ng	-0.01
75) Chrysene-d12	13.79	240	390291	20.00	ng	0.00
86) Perylene-d12	15.16	264	313995	20.00	ng	0.01

System Monitoring Compounds						
5) 2-Fluorophenol	5.24	112	1480434	170.26	ng	0.00
7) Phenol-d6	6.30	99	1617841	152.40	ng	-0.01
23) Nitrobenzene-d5	7.20	82	1280442	112.87	ng	-0.01
41) 2,4,6-Tribromophenol	10.46	330	444821	156.51	ng	-0.01
44) 2-Fluorobiphenyl	9.00	172	1778525	109.89	ng	-0.01
78) Terphenyl-d14	12.73	244	1778443	110.51	ng	-0.01


Target Compounds						Qvalue
10) Phenol	6.31	94	119182	9.85	ng	82
15) Benzyl Alcohol	6.78	79	17664	2.14	ng	# 42
19) n-Nitroso-di-n-propylamine	7.20	70	170858	24.19	ng	# 74
25) Isophorone	7.20	82	1209960	57.81	ng	# 65
37) 2-Methylnaphthalene	8.63	142	6605	Below	Cal	98
49) Dimethylphthalate	9.40	163	374561	17.05	ng	100
56) 2,4-Dinitrotoluene	9.67	165	44255	7.33	ng	# 18
66) 4-Bromophenyl-phenylether	10.46	248	26262	4.32	ng	# 1
70) Phenanthrene	11.17	178	314275	9.93	ng	97
73) Di-n-butylphthalate	11.72	149	7893	Below	Cal	# 78
74) Fluoranthene	12.36	202	525982	16.51	ng	93
76) Benzidine	12.47	184	909	Below	Cal	# 1
77) Pyrene	12.58	202	501219	17.50	ng	97
80) Benzo(a)anthracene	13.77	228	281569	12.78	ng	97
82) Chrysene	13.77	228	281569	12.74	ng	98
85) Indeno(1,2,3-cd)pyrene	16.43	276	179806	9.39	ng	94
87) Benzo(b)fluoranthene	14.78	252	418990	21.43	ng	100
88) Benzo(k)fluoranthene	14.78	252	418990	25.13	ng	99
89) Benzo(a)pyrene	15.10	252	277437	15.99	ng	97
90) Dibenzo(a,h)anthracene	16.43	278	65731	4.61	ng	# 83
91) Benzo(g,h,i)perylene	16.80	276	181765	12.26	ng	# 95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF011017\
Data File : BF092196.D
Acq On : 10 Jan 2017 21:44
Operator : UM/SJ
Sample : I1098-03
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TP3-COMP13

Quant Time: Jan 10 23:36:12 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF122116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 06 15:43:51 2017
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 20.00 ng

RT: 6.62 min Scan# 414

Delta R.T. -0.02 min

Lab File: BF092196.D

Acq: 10 Jan 2017 21:44

Instrument :

BNA_F

ClientSampleId :

TP3-COMP13


Tgt Ion:152 Resp: 145184

Ion Ratio Lower Upper

152 100

150 162.5 124.9 187.3




115 74.9 46.0 69.0#



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#5

2-Fluorophenol

Concen: 170.26 ng

RT: 5.24 min Scan# 293

Delta R.T. 0.00 min

Lab File: BF092196.D

Acq: 10 Jan 2017 21:44


Tgt Ion:112 Resp: 1480434

Ion Ratio Lower Upper

112 100

64 64.9 46.1 69.1



63 33.4 23.8 35.6




Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF0

Ion 63.00 (62.70 to 63.70): BF0





#7

Phenol-d6

Concen: 152.40 ng

RT: 6.30 min Scan# 386

Delta R.T. -0.01 min

Lab File: BF092196.D

Acq: 10 Jan 2017 21:44

Instrument :

BNA_F

ClientSampleId :

TP3-COMP13


Tgt Ion: 99 Resp: 1617841

Ion Ratio Lower Upper

99 100

42 20.1 15.6 23.4




71 36.1 27.5 41.3



Abundance Ion 99.00 (98.70 to 99.70): BF0

Ion 42.00 (41.70 to 42.70): BF0

Ion 71.00 (70.70 to 71.70): BF0



#10

Phenol

Concen: 9.85 ng

RT: 6.31 min Scan# 387

Delta R.T. -0.01 min

Lab File: BF092196.D

Acq: 10 Jan 2017 21:44


Tgt Ion: 94 Resp: 119182

Ion Ratio Lower Upper

94 100

65 31.6 21.4 61.4



66 49.2 44.0 84.0






Abundance Ion 94.00 (93.70 to 94.70): BF0

Ion 65.00 (64.70 to 65.70): BF0

Ion 66.00 (65.70 to 66.70): BF0





#15

Benzyl Alcohol

Concen: 2.14 ng

RT: 6.78 min Scan# 428

Delta R.T. -0.02 min

Lab File: BF092196.D

Acq: 10 Jan 2017 21:44

Instrument :

BNA_F

ClientSampleId :

TP3-COMP13





Tgt Ion: 79 Resp: 17664

Ion Ratio Lower Upper

79 100

108 25.2 61.4 92.0#

77 106.7 50.9 76.3#



#19

n-Nitroso-di-n-propylamine

Concen: 24.19 ng

RT: 7.20 min Scan# 465

Delta R.T. 0.13 min

Lab File: BF092196.D

Acq: 10 Jan 2017 21:44

Tgt Ion: 70 Resp: 170858


Ion Ratio Lower Upper


70 100

42 43.6 49.4 74.0#

101 0.0 7.4 11.2#

130 2.1 14.2 21.4#









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 7.91 min Scan# 527
Delta R.T. -0.02 min
Lab File: BF092196.D
Acq: 10 Jan 2017 21:44




Instrument :
BNA_F
ClientSampleId :
TP3-COMP13


Tgt Ion:	136	Resp:	630791
Ion	Ratio	Lower	Upper
136	100		
137	11.3	8.4	12.6
54	8.3	4.5	6.7#
68	6.4	3.6	5.4#



#23
Nitrobenzene-d5
Concen: 112.87 ng
RT: 7.20 min Scan# 465
Delta R.T. -0.01 min
Lab File: BF092196.D
Acq: 10 Jan 2017 21:44

Tgt Ion:	82	Resp:	1280442
Ion	Ratio	Lower	Upper
82	100		
128	42.8	34.6	52.0
54	47.8	39.5	59.3





#25

Isophorone

Concen: 57.81 ng

RT: 7.20 min Scan# 465

Delta R.T. -0.27 min

Lab File: BF092196.D

Acq: 10 Jan 2017 21:44

Instrument :

BNA_F

ClientSampleId :

TP3-COMP13


Tgt Ion: 82 Resp: 1209960

Ion Ratio Lower Upper

82 100

95 0.0 5.6 8.4#

138 0.0 14.6 22.0#






Abundance Ion 82.00 (81.70 to 82.70): BF092196.D (-463) (-)

Ion 95.00 (94.70 to 95.70): BF092196.D (-463) (-)

Ion 138.00 (137.70 to 138.70): BF092196.D (-463) (-)

Time-->



#38

Acenaphthene-d10

Concen: 20.00 ng

RT: 9.67 min Scan# 681

Delta R.T. -0.01 min

Lab File: BF092196.D

Acq: 10 Jan 2017 21:44


Tgt Ion: 164 Resp: 343425

Ion Ratio Lower Upper

164 100

162 98.8 80.2 120.2

160 43.7 36.9 55.3






Abundance Ion 164.00 (163.70 to 164.70): BF092196.D (-656) (-)

Ion 162.00 (161.70 to 162.70): BF092196.D (-656) (-)

Ion 160.00 (159.70 to 160.70): BF092196.D (-656) (-)

Time-->









#41
 2,4,6-Tribromophenol
 Concen: 156.51 ng
 RT: 10.46 min Scan# 750
 Delta R.T. -0.01 min
 Lab File: BF092196.D
 Acq: 10 Jan 2017 21:44

Instrument :
 BNA_F
 ClientSampleId :
 TP3-COMP13

Tgt Ion:330 Resp: 444821
 Ion Ratio Lower Upper
 330 100
 332 96.0 77.4 116.2
 141 36.0 34.1 51.1




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#44
 2-Fluorobiphenyl
 Concen: 109.89 ng
 RT: 9.00 min Scan# 622
 Delta R.T. -0.01 min
 Lab File: BF092196.D
 Acq: 10 Jan 2017 21:44

Tgt Ion:172 Resp: 1778525
 Ion Ratio Lower Upper
 172 100
 171 35.9 29.4 44.0
 170 24.6 20.2 30.4



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#49
Dimethylphthalate
Concen: 17.05 ng
RT: 9.40 min Scan# 657
Delta R.T. -0.02 min
Lab File: BF092196.D
Acq: 10 Jan 2017 21:44

Instrument :
BNA_F
ClientSampled :
TP3-COMP13

Tgt Ion:163 Resp: 374561
Ion Ratio Lower Upper
163 100
194 3.2 3.0 4.4
164 10.0 8.0 12.0




Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E






#56
2,4-Dinitrotoluene
Concen: 7.33 ng
RT: 9.67 min Scan# 681
Delta R.T. -0.22 min
Lab File: BF092196.D
Acq: 10 Jan 2017 21:44

Tgt Ion:165 Resp: 44255
Ion Ratio Lower Upper
165 100
63 0.0 40.2 60.4#
89 0.0 62.5 93.7#
182 0.0 1.8 2.8#



Abundance Ion 165.00 (164.70 to 165.70): E
Ion 63.00 (62.70 to 63.70): BF0
Ion 89.00 (88.70 to 89.70): BF0
Ion 182.00 (181.70 to 182.70): E









#63
Phenanthrene-d10
Concen: 20.00 ng
RT: 11.14 min Scan# 810
Delta R.T. -0.01 min
Lab File: BF092196.D
Acq: 10 Jan 2017 21:44

Instrument :
BNA_F
ClientSampleId :
TP3-COMP13

Tgt Ion:188 Resp: 583909
Ion Ratio Lower Upper
188 100
94 9.8 10.0 15.0#
80 10.8 11.2 16.8#




Abundance Ion 188.00 (187.70 to 188.70): E
Ion 94.00 (93.70 to 94.70): BF0
Ion 80.00 (79.70 to 80.70): BF0






#66
4-Bromophenyl-phenylether
Concen: 4.32 ng
RT: 10.46 min Scan# 750
Delta R.T. -0.25 min
Lab File: BF092196.D
Acq: 10 Jan 2017 21:44

Tgt Ion:248 Resp: 26262
Ion Ratio Lower Upper
248 100
250 188.7 76.9 115.3#
141 625.6 68.2 102.4#



Abundance Ion 248.00 (247.70 to 248.70): E
Ion 250.00 (249.70 to 250.70): E
Ion 141.00 (140.70 to 141.70): E





#70

Phenanthrene

Concen: 9.93 ng

RT: 11.17 min Scan# 812

Delta R.T. -0.01 min

Lab File: BF092196.D

Acq: 10 Jan 2017 21:44

Instrument :

BNA_F

ClientSampleId :

TP3-COMP13


Tgt Ion:178 Resp: 314275

Ion Ratio Lower Upper

178 100

176 19.1 16.6 25.0




179 15.3 12.8 19.2



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#73

Di-n-butylphthalate

Concen: Below Cal

RT: 11.72 min Scan# 860

Delta R.T. -0.01 min

Lab File: BF092196.D

Acq: 10 Jan 2017 21:44


Tgt Ion:149 Resp: 7893

Ion Ratio Lower Upper

149 100

150 22.1 8.1 12.1#



104 7.3 4.6 7.0#




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E





#74

Fluoranthene

Concen: 16.51 ng

RT: 12.36 min Scan# 916

Delta R.T. -0.01 min

Lab File: BF092196.D

Acq: 10 Jan 2017 21:44

Instrument :

BNA_F

ClientSampleId :

TP3-COMP13


Tgt Ion:202 Resp: 525982

Ion Ratio Lower Upper

202 100

101 9.6 0.0 34.6




203 17.3 0.0 38.7



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#75

Chrysene-d12

Concen: 20.00 ng

RT: 13.79 min Scan# 1041

Delta R.T. 0.00 min

Lab File: BF092196.D

Acq: 10 Jan 2017 21:44


Tgt Ion:240 Resp: 390291

Ion Ratio Lower Upper

240 100

120 7.0 12.9 19.3#



236 25.3 20.9 31.3




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E





#77

Pyrene

Concen: 17.50 ng

RT: 12.58 min Scan# 936

Delta R.T. 0.00 min

Lab File: BF092196.D

Acq: 10 Jan 2017 21:44

Instrument :

BNA_F

ClientSampleId :

TP3-COMP13


Tgt Ion:202 Resp: 501219

Ion Ratio Lower Upper

202 100

200 20.5 17.7 26.5




203 17.7 14.7 22.1



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E



#78

Terphenyl-d14

Concen: 110.51 ng

RT: 12.73 min Scan# 949

Delta R.T. -0.01 min

Lab File: BF092196.D

Acq: 10 Jan 2017 21:44


Tgt Ion:244 Resp: 1778443

Ion Ratio Lower Upper

244 100

212 7.6 6.2 9.2



122 14.6 11.4 17.2




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E





#80

Benzo(a)anthracene

Concen: 12.78 ng

RT: 13.77 min Scan# 1040

Delta R.T. 0.00 min

Lab File: BF092196.D

Acq: 10 Jan 2017 21:44

Instrument :

BNA_F

ClientSampleId :

TP3-COMP13


Tgt Ion:228 Resp: 281569

Ion Ratio Lower Upper

228 100

226 30.2 22.4 33.6




229 20.8 16.2 24.4



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#82

Chrysene

Concen: 12.74 ng

RT: 13.77 min Scan# 1040

Delta R.T. -0.03 min

Lab File: BF092196.D

Acq: 10 Jan 2017 21:44


Tgt Ion:228 Resp: 281569

Ion Ratio Lower Upper

228 100

226 30.2 25.4 38.0



229 20.8 16.2 24.2




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E





#85

Indeno(1,2,3-cd)pyrene

Concen: 9.39 ng

RT: 16.43 min Scan# 1272

Delta R.T. 0.00 min

Lab File: BF092196.D

Acq: 10 Jan 2017 21:44

Instrument :

BNA_F

ClientSampleId :

TP3-COMP13


Tgt Ion: 276 Resp: 179806

Ion Ratio Lower Upper

276 100

138 25.2 21.8 32.6




277 29.8 20.2 30.4



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E



#86

Perylene-d12

Concen: 20.00 ng

RT: 15.16 min Scan# 1161

Delta R.T. 0.01 min

Lab File: BF092196.D

Acq: 10 Jan 2017 21:44


Tgt Ion: 264 Resp: 313995

Ion Ratio Lower Upper

264 100

260 23.2 19.2 28.8



265 23.0 17.4 26.0




Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E









#87
Benzo(b)fluoranthene
Concen: 21.43 ng
RT: 14.78 min Scan# 1128
Delta R.T. 0.00 min
Lab File: BF092196.D
Acq: 10 Jan 2017 21:44




Instrument :
BNA_F
ClientSampleId :
TP3-COMP13


Tgt Ion:252 Resp: 418990
Ion Ratio Lower Upper
252 100
253 22.7 18.2 27.4
125 12.1 9.8 14.6



#88
Benzo(k)fluoranthene
Concen: 25.13 ng
RT: 14.78 min Scan# 1128
Delta R.T. -0.02 min
Lab File: BF092196.D
Acq: 10 Jan 2017 21:44

Tgt Ion:252 Resp: 418990
Ion Ratio Lower Upper
252 100
253 22.7 18.5 27.7
125 12.1 8.9 13.3





#89

Benzo(a)pyrene

Concen: 15.99 ng

RT: 15.10 min Scan# 1156

Delta R.T. 0.00 min

Lab File: BF092196.D

Acq: 10 Jan 2017 21:44


Instrument :

BNA_F




ClientSampleId :

TP3-COMP13

Tgt Ion:	252	Resp:	277437
Ion Ratio		Lower	Upper
252	100		
253	23.3	18.2	27.2
125	10.4	10.0	15.0



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E



#90

Dibenzo(a,h)anthracene

Concen: 4.61 ng


RT: 16.43 min Scan# 1272

Delta R.T. -0.01 min



Lab File: BF092196.D


Acq: 10 Jan 2017 21:44

Tgt Ion:	278	Resp:	65731
Ion Ratio		Lower	Upper
278	100		
139	26.6	14.8	22.2#
279	32.9	19.8	29.6#



Abundance Ion 278.00 (277.70 to 278.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 279.00 (278.70 to 279.70): E





#91

Benzo(g,h,i)perylene

Concen: 12.26 ng

RT: 16.80 min Scan# 1305

Delta R.T. 0.00 min

Lab File: BF092196.D


Acq: 10 Jan 2017 21:44

Instrument :

BNA_F

ClientSampleId :

TP3-COMP13




Tgt Ion: 276 Resp: 181765

Ion Ratio Lower Upper

276 100

277 24.9 19.2 28.8

138 24.3 16.0 24.0#



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

