

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF011022\
 Data File : BF126562.D
 Acq On : 10 Jan 2022 13:28
 Operator : JU/CG
 Sample : N1065-06
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 AMI-2

Quant Time: Jan 10 17:15:55 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 10 14:50:31 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 01/11/2022
 Supervised By :Jagrut Upadhyay 01/11/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.769	152	96859	20.000	ng	0.00
21) Naphthalene-d8	8.051	136	417224	20.000	ng	# 0.00
39) Acenaphthene-d10	9.810	164	204094	20.000	ng	0.00
64) Phenanthrene-d10	11.298	188	332098	20.000	ng	# 0.00
76) Chrysene-d12	13.939	240	193952	20.000	ng	0.00
86) Perylene-d12	15.392	264	165353	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.386	112	563201	83.502	ng	0.00
7) Phenol-d6	6.404	99	795781	91.093	ng	0.00
23) Nitrobenzene-d5	7.333	82	510782	62.860	ng	0.00
42) 2,4,6-Tribromophenol	10.598	330	133506	84.651	ng	0.00
45) 2-Fluorobiphenyl	9.133	172	784664	67.655	ng	0.00
79) Terphenyl-d14	12.886	244	694974	69.620	ng	0.00
Target Compounds						
						Qvalue
75) Fluoranthene	12.510	202	54431	3.513	ng	97
78) Pyrene	12.733	202	55827	3.406	ng	98
83) Chrysene	13.962	228	23775	2.078	ng	99
88) Benzo(b)fluoranthene	14.980	252	27847m	2.827	ng	
90) Benzo(a)pyrene	15.333	252	17820	2.158	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF011022\
Data File : BF126562.D
Acq On : 10 Jan 2022 13:28
Operator : JU/CG
Sample : N1065-06
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
AMI-2

Quant Time: Jan 10 17:15:55 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 10 14:50:31 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 01/11/2022
Supervised By :Jagrut Upadhyay 01/11/2022

