




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF011023\
Data File : BF131995.D
Acq On : 10 Jan 2023 14:33
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 11 04:31:12 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF010923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 10 00:32:25 2023
Response via : Initial Calibration






TIC: BF131995.D\data.ms

(70) Pentachlorophenol (C)**11.133min (-0.006) 20586.78 ng****response 103859**

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	65.90	64.63
264.00	64.50	60.61
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF011023\
 Data File : BF131995.D
 Acq On : 10 Jan 2023 14:33
 Operator : CG\JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 11 04:31:12 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF010923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 10 00:32:25 2023
 Response via : Initial Calibration



TIC: BF131995.D\data.ms

(77) Benzidine

12.680min (+ 0.005) 0.00 ng

response 679404

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	16.00	13.72
183.00	11.20	11.58
0.00	0.00	0.00