

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF011116\
 Data File : BF084375.D
 Acq On : 11 Jan 2016 13:06
 Operator : SJ/UM
 Sample : H1043-09 5X
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-5S(0-10)

Quant Time: Jan 12 03:02:21 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF123115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 04 12:22:40 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.85	152	249478	20.00	ng	-0.05
21) Naphthalene-d8	8.14	136	1007311	20.00	ng	-0.05
38) Acenaphthene-d10	9.89	164	447769	20.00	ng	-0.06
63) Phenanthrene-d10	11.38	188	768773	20.00	ng	-0.05
75) Chrysene-d12	14.02	240	480758	20.00	ng	-0.05
86) Perylene-d12	15.45	264	437608	20.00	ng	-0.06
System Monitoring Compounds						
5) 2-Fluorophenol	5.46	112	450380	29.20	ng	-0.03
7) Phenol-d6	6.49	99	537099	27.73	ng	-0.05
23) Nitrobenzene-d5	7.41	82	347822	19.22	ng	-0.06
41) 2,4,6-Tribromophenol	10.68	330	126134	27.90	ng	-0.06
44) 2-Fluorobiphenyl	9.22	172	628772	19.85	ng	-0.05
78) Terphenyl-d14	12.96	244	412113	19.13	ng	-0.06
Target Compounds						Qvalue
49) Dimethylphthalate	9.61	163	145689	4.55	ng	# 90
70) Phenanthrene	11.40	178	168658	4.19	ng	99
74) Fluoranthene	12.59	202	217967	5.19	ng	96
77) Pyrene	12.82	202	168505	4.47	ng	99
82) Chrysene	14.04	228	66872m	2.47	ng	
87) Benzo(b)fluoranthene	15.04	252	85303m	2.98	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF011116\
Data File : BF084375.D
Acq On : 11 Jan 2016 13:06
Operator : SJ/UM
Sample : H1043-09 5X
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-5S(0-10)

Quant Time: Jan 12 03:02:21 2016
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF123115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 04 12:22:40 2016
Response via : Initial Calibration

