

Data Path : U:\HPCHEM1\BNA F\Data\BF011118\
 Data File : BF102018.D
 Acq On : 11 Jan 2018 15:05
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 11 15:36:48 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF010918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 09 15:20:36 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	85	0.00
2	1,4-Dioxane	0.707	0.659	6.8	80	0.00
3	Pyridine	1.930	1.825	5.4	79	0.00
4	n-Nitrosodimethylamine	0.808	0.734	9.2	76	0.00
5 S	2-Fluorophenol	1.416	1.314	7.2	80	0.00
6	Aniline	2.381	2.237	6.0	79	0.00
7 S	Phenol-d6	1.690	1.586	6.2	81	0.00
8	2-Chlorophenol	1.421	1.370	3.6	81	0.00
9	Benzaldehyde	1.173	1.078	8.1	77	0.00
10 C	Phenol	1.910	1.767	7.5	77	0.00
11	bis(2-Chloroethyl)ether	1.454	1.354	6.9	80	0.00
12	1,3-Dichlorobenzene	1.638	1.550	5.4	81	0.00
13 C	1,4-Dichlorobenzene	1.634	1.575	3.6	83	0.00
14	1,2-Dichlorobenzene	1.512	1.449	4.2	82	0.00
15	Benzyl Alcohol	1.236	1.169	5.4	80	0.00
16	2,2'-oxybis(1-Chloropropane	2.680	2.491	7.1	78	0.00
17	2-Methylphenol	1.281	1.230	4.0	81	0.00
18	Hexachloroethane	0.575	0.565	1.7	83	0.00
19 P	n-Nitroso-di-n-propylamine	1.107	1.004	9.3	78	0.00
20	3+4-Methylphenols	1.539	1.415	8.1	79	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	85	0.00
22	Acetophenone	0.482	0.439	8.9	80	0.00
23 S	Nitrobenzene-d5	0.340	0.340	0.0	83	0.00
24	Nitrobenzene	0.366	0.358	2.2	82	0.00
25	Isophorone	0.672	0.621	7.6	80	0.00
26 C	2-Nitrophenol	0.153	0.180	-17.6	95	0.00
27	2,4-Dimethylphenol	0.286	0.273	4.5	82	0.00
28	bis(2-Chloroethoxy)methane	0.406	0.377	7.1	81	0.00
29 C	2,4-Dichlorophenol	0.271	0.263	3.0	82	0.00
30	1,2,4-Trichlorobenzene	0.284	0.271	4.6	83	0.00
31	Naphthalene	0.999	0.935	6.4	82	0.00
32	Benzoic acid	0.217	0.179	17.5	64	0.00
33	4-Chloroaniline	0.427	0.408	4.4	82	0.00
34 C	Hexachlorobutadiene	0.150	0.145	3.3	83	0.00
35	Caprolactam	0.093	0.093	0.0	83	0.00
36 C	4-Chloro-3-methylphenol	0.297	0.288	3.0	83	0.00
37	2-Methylnaphthalene	0.658	0.617	6.2	82	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	85	0.00
39	1,2,4,5-Tetrachlorobenzene	0.566	0.539	4.8	83	0.00
40 P	Hexachlorocyclopentadiene	0.309	0.325	-5.2	87	0.00
41 S	2,4,6-Tribromophenol	0.175	0.181	-3.4	88	0.00
42 C	2,4,6-Trichlorophenol	0.390	0.397	-1.8	86	0.00
43	2,4,5-Trichlorophenol	0.441	0.436	1.1	83	0.00
44 S	2-Fluorobiphenyl	1.280	1.253	2.1	82	0.00

Data Path : U:\HPCHEM1\BNA F\Data\BF011118\
 Data File : BF102018.D
 Acq On : 11 Jan 2018 15:05
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 11 15:36:48 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF010918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 09 15:20:36 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.768	1.654	6.4	81	0.00
46	2-Chloronaphthalene	1.371	1.299	5.3	82	0.00
47	2-Nitroaniline	0.411	0.447	-8.8	86	0.00
48	Acenaphthylene	2.176	2.058	5.4	82	0.00
49	Dimethylphthalate	1.604	1.541	3.9	83	0.00
50	2,6-Dinitrotoluene	0.320	0.348	-8.7	86	0.00
51 C	Acenaphthene	1.320	1.208	8.5	80	0.00
52	3-Nitroaniline	0.404	0.421	-4.2	84	0.00
53 P	2,4-Dinitrophenol	0.098	0.153	-56.1#	127	0.00
54	Dibenzofuran	1.812	1.708	5.7	82	0.00
55 P	4-Nitrophenol	0.301	0.322	-7.0	85	0.00
56	2,4-Dinitrotoluene	0.402	0.456	-13.4	89	0.00
57	Fluorene	1.253	1.257	-0.3	85	0.00
58	2,3,4,6-Tetrachlorophenol	0.314	0.314	0.0	84	0.00
59	Diethylphthalate	1.596	1.500	6.0	82	0.00
60	4-Chlorophenyl-phenylether	0.533	0.534	-0.2	85	0.00
61	4-Nitroaniline	0.384	0.419	-9.1	88	0.00
62	Azobenzene	1.588	1.472	7.3	81	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	88	0.00
64	4,6-Dinitro-2-methylphenol	0.098	0.134	-36.7#	110	0.00
65 c	n-Nitrosodiphenylamine	0.749	0.687	8.3	83	0.00
66	4-Bromophenyl-phenylether	0.211	0.201	4.7	84	0.00
67	Hexachlorobenzene	0.231	0.225	2.6	86	0.00
68	Atrazine	0.216	0.214	0.9	86	0.00
69 C	Pentachlorophenol	0.141	0.147	-4.3	85	0.00
70	Phenanthrene	1.168	1.079	7.6	84	0.00
71	Anthracene	1.185	1.110	6.3	85	0.00
72	Carbazole	1.165	1.099	5.7	85	0.00
73	Di-n-butylphthalate	1.429	1.343	6.0	83	0.00
74 C	Fluoranthene	1.149	1.098	4.4	87	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
76	Benzidine	0.966	0.804	16.8	79	0.00
77	Pyrene	1.768	1.602	9.4	87	0.00
78 S	Terphenyl-d14	0.963	0.847	12.0	87	0.00
79	Butylbenzylphthalate	0.811	0.802	1.1	90	0.00
80	Benzo(a)anthracene	1.276	1.149	10.0	89	0.00
81	3,3'-Dichlorobenzidine	0.472	0.467	1.1	94	0.00
82	Chrysene	1.331	1.226	7.9	90	0.00
83	Bis(2-ethylhexyl)phthalate	0.931	0.908	2.5	93	0.00
84 c	Di-n-octyl phthalate	1.544	1.607	-4.1	98	0.00
85	Indeno(1,2,3-cd)pyrene	0.968	0.883	8.8	92	0.00
86 I	Perylene-d12	1.000	1.000	0.0	101	0.00
87	Benzo(b)fluoranthene	1.304	1.275	2.2	93	0.00

Data Path : U:\HPCHEM1\BNA F\Data\BF011118\
Data File : BF102018.D
Acq On : 11 Jan 2018 15:05
Operator : SJ/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jan 11 15:36:48 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF010918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 09 15:20:36 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.258	1.197	4.8	99	0.00
89 C	Benzo(a)pyrene	1.173	1.133	3.4	97	0.00
90	Dibenzo(a,h)anthracene	0.936	0.862	7.9	91	0.00
91	Benzo(a,h,i)perylene	0.943	0.870	7.7	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0