




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF011123\
 Data File : BF132017.D
 Acq On : 11 Jan 2023 10:14
 Operator : CG\JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 12 03:29:50 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF010923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 10 00:32:25 2023
 Response via : Initial Calibration






TIC: BF132017.D\data.ms

(70) Pentachlorophenol (C)**11.133min (-0.006) 17874.41 ng****response 116191**

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	65.90	64.32
264.00	64.50	61.80
0.00	0.00	0.00

Quant Time: Jan 12 03:29:50 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF010923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 10 00:32:25 2023
Response via : Initial Calibration



TIC: BF132017.D\data.ms

(77) Benzidine

12.680min (+ 0.005) 0.00 ng

response 900922

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	16.00	14.22
183.00	11.20	11.19
0.00	0.00	0.00