


Data Path : Z:\HPCHEM1\BNA F\Data\BF011216\
 Data File : BF084411.D
 Acq On : 12 Jan 2016 10:00
 Operator : SJ/UM
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF123115.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jan 14 14:15:52 2016



AutoFind: Scans 834, 835, 836; Background Corrected with Scan 831

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	44.0	68786	PASS
68	69	0.00	2	1.7	1063	PASS
69	198	0.00	100	40.1	62583	PASS
70	69	0.00	2	0.7	439	PASS
127	198	10	80	52.2	81543	PASS
197	198	0.00	2	0.4	578	PASS
198	198	100	100	100.0	156176	PASS
199	198	5	9	7.2	11266	PASS
275	198	10	60	22.8	35641	PASS
365	198	1	100	2.7	4265	PASS
441	442	0.01	24	16.3	20992	PASS
442	198	50	100	82.6	128991	PASS
443	442	15	24	18.9	24392	PASS