

Data Path : Z:\HPCHEM1\BNA F\DATA\BF011217\
 Data File : BF092206.D
 Acq On : 12 Jan 2017 14:20
 Operator : UM/SJ
 Sample : I1029-03
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 COMP-002

Quant Time: Jan 13 13:44:22 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF122116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 12 16:14:19 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.60	152	200587	20.00	ng	0.00
21) Naphthalene-d8	7.88	136	829152	20.00	ng	0.00
38) Acenaphthene-d10	9.62	164	365623	20.00	ng	0.00
63) Phenanthrene-d10	11.11	188	604770	20.00	ng	0.00
75) Chrysene-d12	13.73	240	449775	20.00	ng	0.00
86) Perylene-d12	15.09	264	335493	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.20	112	1838347	153.03	ng	0.00
7) Phenol-d6	6.26	99	2103401	143.41	ng	0.00
23) Nitrobenzene-d5	7.17	82	1416024	94.96	ng	0.00
41) 2,4,6-Tribromophenol	10.42	330	449076	148.42	ng	0.00
44) 2-Fluorobiphenyl	8.96	172	2107161	124.00	ng	0.00
78) Terphenyl-d14	12.70	244	1932543	104.21	ng	0.00
Target Compounds						
10) Phenol	6.28	94	148478	8.88	ng	78
49) Dimethylphthalate	9.36	163	264360	11.30	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF011217\
Data File : BF092206.D
Acq On : 12 Jan 2017 14:20
Operator : UM/SJ
Sample : I1029-03
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
COMP-002

Quant Time: Jan 13 13:44:22 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF122116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 12 16:14:19 2017
Response via : Initial Calibration

