

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_F\DATA\BF011218\
 Data File : BF102061.D
 Acq On : 12 Jan 2018 23:34
 Operator : SJ/JU
 Sample : J1106-04
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 CLEAN-SAND-16

Quant Time: Jan 13 01:21:20 2018
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF010918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 11 12:53:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.73	152	436463	20.00	ng	0.00
21) Naphthalene-d8	8.02	136	1703024	20.00	ng	0.00
38) Acenaphthene-d10	9.77	164	692036	20.00	ng	0.00
63) Phenanthrene-d10	11.25	188	992076	20.00	ng	0.00
75) Chrysene-d12	13.89	240	711020	20.00	ng	0.00
86) Perylene-d12	15.31	264	538274	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.35	112	1774420	57.41	ng	0.02
7) Phenol-d6	6.37	99	2103737	57.05	ng	0.01
23) Nitrobenzene-d5	7.30	82	1347182	46.49	ng	0.00
41) 2,4,6-Tribromophenol	10.56	330	333134	55.16	ng	0.00
44) 2-Fluorobiphenyl	9.09	172	2050801	46.30	ng	0.00
78) Terphenyl-d14	12.84	244	1385297	40.45	ng	0.00
Target Compounds						
10) Phenol	6.38	94	103512	2.484	ng	76
49) Dimethylphthalate	9.49	163	247942	4.466	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA_F\DATA\BF011218\
Data File : BF102061.D
Acq On : 12 Jan 2018 23:34
Operator : SJ/JU
Sample : J1106-04
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
CLEAN-SAND-16

Quant Time: Jan 13 01:21:20 2018
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF010918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 11 12:53:54 2018
Response via : Initial Calibration

