


Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF011224\  
Data File : BF137043.D  
Acq On : 12 Jan 2024 12:31  
Operator : CG\JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF122023.M  
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
Last Update : Thu Dec 21 01:54:25 2023






AutoFind: Scans 1611, 1612, 1613; Background Corrected with Scan 1606

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 38.6         | 52203      | PASS                |
| 68             | 69              | 0.00            | 2               | 1.5          | 751        | PASS                |
| 69             | 198             | 0.00            | 100             | 35.9         | 48579      | PASS                |
| 70             | 69              | 0.00            | 2               | 0.5          | 251        | PASS                |
| 127            | 198             | 10              | 80              | 48.2         | 65189      | PASS                |
| 197            | 198             | 0.00            | 2               | 0.0          | 0          | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 135237     | PASS                |
| 199            | 198             | 5               | 9               | 6.9          | 9285       | PASS                |
| 275            | 198             | 10              | 60              | 28.5         | 38504      | PASS                |
| 365            | 198             | 1               | 100             | 3.4          | 4664       | PASS                |
| 441            | 198             | 0.01            | 100             | 15.6         | 21108      | PASS                |
| 442            | 442             | 50              | 100             | 100.0        | 131677     | PASS                |
| 443            | 442             | 15              | 24              | 19.0         | 24987      | PASS                |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF011224\  
Data File : BF137043.D  
Acq On : 12 Jan 2024 12:31  
Operator : CG\JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
DFTPP

Quant Time: Jan 12 14:57:47 2024  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF122023.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Dec 21 01:54:25 2023  
Response via : Initial Calibration



TIC: BF137043.D\data.ms

**(70) Pentachlorophenol (C)**

**11.128min (+ 0.006) 80510.76 ng**




**response 116112**

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 265.70 | 100.00 | 100.00 |
| 268.00 | 65.70  | 66.32  |
| 264.00 | 61.20  | 62.30  |
| 0.00   | 0.00   | 0.00   |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF011224\  
Data File : BF137043.D  
Acq On : 12 Jan 2024 12:31  
Operator : CG\JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
DFTPP

Quant Time: Jan 12 14:57:47 2024  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_F\Methods\8270-BF122023.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Dec 21 01:54:25 2023  
Response via : Initial Calibration



TIC: BF137043.D\data.ms

(77) Benzidine

12.669min (+ 0.006) 310810.36 ng

response 687335

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 184.00 | 100.00 | 100.00 |
| 185.00 | 14.00  | 13.96  |
| 183.00 | 11.60  | 11.02  |
| 0.00   | 0.00   | 0.00   |

### DDT Breakdown

| Date          | Instrument Name  | DFTPP Data File    |
|---------------|------------------|--------------------|
| 1/12/2024     | BNA_F            | <u>BF137043.D</u>  |
|               |                  |                    |
|               |                  |                    |
| Compound Name | Response         | Retention Time     |
|               |                  |                    |
| DDT           | 335771           | 13.486             |
| DDD           | 9215             | 13.186             |
| DDE           | 539              | 12.845             |
|               |                  |                    |
|               |                  |                    |
| SUM(DDD+DDE)  | SUM(DDT+DDD+DDE) | % Breakdown Of DDT |
|               |                  |                    |
| 9754          | 345525           | 2.82               |
|               |                  |                    |

Instrument :  
BNA\_F  
ClientSampleId :  
DFTPP