


Data Path : Z:\HPCHEM1\BNA_F\Data\BF011315\
 Data File : BF076839.D
 Acq On : 13 Jan 2015 11:40
 Operator : IZ / TP
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF010915.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jan 12 10:43:48 2015



AutoFind: Scans 1514, 1515, 1516; Background Corrected with Scan 1510

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.7	4213	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	40.8	4564	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	52.9	5916	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	11180	PASS
199	198	5	9	6.7	744	PASS
275	198	10	60	24.1	2691	PASS
365	198	1	100	2.1	239	PASS
441	442	0.01	24	15.5	1573	PASS
442	198	50	100	90.6	10129	PASS
443	442	15	24	19.0	1929	PASS