

Data Path : Z:\HPCHEM1\BNA F\DATA\BF011317\
 Data File : BF092245.D
 Acq On : 13 Jan 2017 13:18
 Operator : UM/SJ
 Sample : I1131-06
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SW-003

Manual Integrations
 APPROVED

Sohil
 1/16/2017 12:00:38 PM

Quant Time: Jan 16 09:59:39 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF122116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 13 12:48:58 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.59	152	189687	20.00	ng	0.00
21) Naphthalene-d8	7.87	136	729832	20.00	ng	-0.01
38) Acenaphthene-d10	9.62	164	322703	20.00	ng	-0.01
63) Phenanthrene-d10	11.10	188	450626	20.00	ng	0.00
75) Chrysene-d12	13.72	240	325279	20.00	ng	-0.01
86) Perylene-d12	15.08	264	310106	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.19	112	1962360	172.74	ng	0.02
7) Phenol-d6	6.26	99	2157786	155.57	ng	0.00
23) Nitrobenzene-d5	7.16	82	1266662	96.51	ng	0.00
41) 2,4,6-Tribromophenol	10.42	330	393647	147.40	ng	0.00
44) 2-Fluorobiphenyl	8.95	172	1957662	131.33	ng	0.00
78) Terphenyl-d14	12.68	244	1146125	85.45	ng	-0.01
Target Compounds						
10) Phenol	6.27	94	44707	2.83	ng	92
49) Dimethylphthalate	9.35	163	450964	21.85	ng	99
70) Phenanthrene	11.12	178	60873	2.49	ng	97
77) Pyrene	12.53	202	67282	2.82	ng	97
87) Benzo(b)fluoranthene	14.71	252	42095m	2.18	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF011317\
Data File : BF092245.D
Acq On : 13 Jan 2017 13:18
Operator : UM/SJ
Sample : I1131-06
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SW-003

Manual Integrations
APPROVED

Sohil
1/16/2017 12:00:38 PM

Quant Time: Jan 16 09:59:39 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF122116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 13 12:48:58 2017
Response via : Initial Calibration

