




Data Path : Z:\HPCHEM1\BNA_F\DATA\BF011317\
Data File : BF092249.D
Acq On : 13 Jan 2017 17:41
Operator : UM/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 16 07:32:37 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF122116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 16 01:02:36 2017
Response via : Initial Calibration



TIC: BF092249.D

(69) Pentachlorophenol (C)




10.908min (-0.011) 0.00ng

response 143995

Ion	Exp%	Act%
265.70	100	100
268.00	66.60	66.25
264.00	62.90	63.40
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF011317\
Data File : BF092249.D
Acq On : 13 Jan 2017 17:41
Operator : UM/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 16 07:32:37 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF122116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 16 01:02:36 2017
Response via : Initial Calibration



TIC: BF092249.D

(76) Benzidine

12.428min (-0.011) 0.00ng

response 1155225

Ion	Exp%	Act%
184.00	100	100
185.00	16.70	14.54
183.00	11.50	11.62
0.00	0.00	0.00