


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF011317\
 Data File : BF092238.D
 Acq On : 13 Jan 2017 9:51
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF122116.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Jan 06 15:43:51 2017



AutoFind: Scans 838, 839, 840; Background Corrected with Scan 835

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.1	40157	PASS
68	69	0.00	2	1.4	555	PASS
69	198	0.00	100	38.9	39899	PASS
70	69	0.00	2	0.5	197	PASS
127	198	10	80	48.6	49848	PASS
197	198	0.00	2	0.6	594	PASS
198	198	100	100	100.0	102662	PASS
199	198	5	9	6.6	6817	PASS
275	198	10	60	24.7	25343	PASS
365	198	1	100	2.9	3022	PASS
441	442	0.01	24	15.7	15372	PASS
442	198	50	100	95.2	97697	PASS
443	442	15	24	18.8	18344	PASS