


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF011317\
 Data File : BF092272.D
 Acq On : 14 Jan 2017 8:42
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF122116.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Jan 06 15:43:51 2017



AutoFind: Scans 837, 838, 839; Background Corrected with Scan 834

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.7	44758	PASS
68	69	0.00	2	1.3	585	PASS
69	198	0.00	100	38.6	44647	PASS
70	69	0.00	2	0.3	139	PASS
127	198	10	80	50.1	57887	PASS
197	198	0.00	2	0.5	588	PASS
198	198	100	100	100.0	115581	PASS
199	198	5	9	6.8	7849	PASS
275	198	10	60	23.6	27318	PASS
365	198	1	100	2.7	3094	PASS
441	442	0.01	24	16.0	17138	PASS
442	198	50	100	92.5	106865	PASS
443	442	15	24	19.2	20496	PASS