

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF011320\
 Data File : BF118534.D
 Acq On : 13 Jan 2020 16:51
 Operator : JU
 Sample : L1106-06
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 CIY1-SB-01-0-52

Quant Time: Jan 14 11:18:45 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF010820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 14 11:02:16 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.82	152	55285	20.00	ng	0.00
21) Naphthalene-d8	8.10	136	219410	20.00	ng	0.00
39) Acenaphthene-d10	9.86	164	120137	20.00	ng	-0.01
64) Phenanthrene-d10	11.35	188	252512	20.00	ng	0.00
76) Chrysene-d12	14.00	240	212272	20.00	ng	-0.01
87) Perylene-d12	15.48	264	194746	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.43	112	416495	124.36	ng	0.00
7) Phenol-d6	6.46	99	549578	137.26	ng	0.00
23) Nitrobenzene-d5	7.38	82	336601	80.59	ng	-0.01
42) 2,4,6-Tribromophenol	10.66	330	204289	144.11	ng	0.00
45) 2-Fluorobiphenyl	9.17	172	626870	74.84	ng	-0.01
79) Terphenyl-d14	12.94	244	899565	67.16	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.57	163	53932	5.849	ng	98
71) Phenanthrene	11.37	178	33866	2.394	ng	100
75) Fluoranthene	12.57	202	60498	4.208	ng	98
78) Pyrene	12.80	202	63752	3.650	ng	97
80) Butylbenzylphthalate	13.42	149	330373	41.305	ng	90
83) Chrysene	14.03	228	30064	2.082	ng	97
84) Bis(2-ethylhexyl)phthalate	13.97	149	94817	8.731	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF011320\
Data File : BF118534.D
Acq On : 13 Jan 2020 16:51
Operator : JU
Sample : L1106-06
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
CIY1-SB-01-0-52

Quant Time: Jan 14 11:18:45 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF010820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jan 14 11:02:16 2020
Response via : Initial Calibration

