

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF011323\
 Data File : BF132057.D
 Acq On : 13 Jan 2023 13:50
 Operator : CG\JU
 Sample : PB150221BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB150221BL

Quant Time: Jan 16 01:16:16 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF010923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 12 23:59:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.775	152	79988	20.000	ng	0.00
21) Naphthalene-d8	8.063	136	303000	20.000	ng	0.00
39) Acenaphthene-d10	9.822	164	185093	20.000	ng	0.00
64) Phenanthrene-d10	11.316	188	364642	20.000	ng	0.00
76) Chrysene-d12	13.962	240	274503	20.000	ng	# 0.00
86) Perylene-d12	15.427	264	196039	20.000	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.404	112	685824	141.282	ng	0.02
7) Phenol-d6	6.428	99	863412	134.149	ng	0.00
23) Nitrobenzene-d5	7.351	82	614093	84.115	ng	0.00
42) 2,4,6-Tribromophenol	10.622	330	271000	127.340	ng	0.00
45) 2-Fluorobiphenyl	9.145	172	1124907	82.044	ng	0.00
79) Terphenyl-d14	12.910	244	1499937	78.317	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF011323\
Data File : BF132057.D
Acq On : 13 Jan 2023 13:50
Operator : CG\JU
Sample : PB150221BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB150221BL

Quant Time: Jan 16 01:16:16 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF010923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 12 23:59:08 2023
Response via : Initial Calibration

