




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF011325\
Data File : BF141122.D
Acq On : 13 Jan 2025 10:17
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 13 11:42:36 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 10 22:54:19 2025
Response via : Initial Calibration



TIC: BF141122.D\data.ms

(70) Pentachlorophenol (C)




11.145min (-0.000) 73211.24 ng

response 203698

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.00	63.53
264.00	62.70	62.96
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF011325\
Data File : BF141122.D
Acq On : 13 Jan 2025 10:17
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 13 11:42:36 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 10 22:54:19 2025
Response via : Initial Calibration



TIC: BF141122.D\data.ms

(77) Benzidine

12.686min (+ 0.006) 792897.58 ng m

response 926339

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.20	14.06
183.00	11.50	11.93
0.00	0.00	0.00