


Data Path : Z:\HPCHEM1\BNA F\Data\BF011515\
 Data File : BF076874.D
 Acq On : 15 Jan 2015 9:18
 Operator : IZ / TP
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF011415.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jan 15 11:52:04 2015



AutoFind: Scans 1513, 1514, 1515; Background Corrected with Scan 1510

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	44.7	4575	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	44.0	4504	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	55.7	5701	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	10231	PASS
199	198	5	9	6.2	639	PASS
275	198	10	60	24.6	2519	PASS
365	198	1	100	1.5	150	PASS
441	442	0.01	24	16.7	1569	PASS
442	198	50	100	91.6	9371	PASS
443	442	15	24	18.8	1760	PASS