

Data Path : Z:\HPCHEM1\BNA F\DATA\BF011516\
 Data File : BF084516.D
 Acq On : 16 Jan 2016 3:29
 Operator : SJ/UM
 Sample : H1144-02
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-2-16BGS

Quant Time: Jan 16 04:10:26 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF123115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jan 16 03:20:07 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.82	152	245596	20.00	ng	0.00
21) Naphthalene-d8	8.11	136	1021527	20.00	ng	0.00
38) Acenaphthene-d10	9.86	164	452062	20.00	ng	-0.01
63) Phenanthrene-d10	11.34	188	791644	20.00	ng	0.00
75) Chrysene-d12	13.99	240	616018	20.00	ng	0.00
86) Perylene-d12	15.40	264	538483	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.44	112	1926705	126.89	ng	0.01
7) Phenol-d6	6.46	99	2213760	116.09	ng	-0.01
23) Nitrobenzene-d5	7.39	82	1600206	87.18	ng	0.00
41) 2,4,6-Tribromophenol	10.66	330	593048	129.94	ng	0.00
44) 2-Fluorobiphenyl	9.18	172	2181704	84.89	ng	-0.01
78) Terphenyl-d14	12.93	244	1776069	64.36	ng	0.00
Target Compounds						
49) Dimethylphthalate	9.58	163	656198	20.30	ng	Qvalue # 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF011516\
Data File : BF084516.D
Acq On : 16 Jan 2016 3:29
Operator : SJ/UM
Sample : H1144-02
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-2-16BGS

Quant Time: Jan 16 04:10:26 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF123115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jan 16 03:20:07 2016
Response via : Initial Calibration

