


Data Path : Z:\HPCHEM1\BNA F\Data\BF011615\
 Data File : BF076908.D
 Acq On : 16 Jan 2015 11:35
 Operator : IZ / TP
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF011415.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jan 15 11:52:04 2015



AutoFind: Scans 1512, 1513, 1514; Background Corrected with Scan 1509

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.9	5728	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	38.9	5578	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	52.7	7557	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	14340	PASS
199	198	5	9	5.9	849	PASS
275	198	10	60	21.4	3063	PASS
365	198	1	100	1.4	199	PASS
441	442	0.01	24	14.9	1781	PASS
442	198	50	100	83.2	11924	PASS
443	442	15	24	20.8	2486	PASS