

Data Path : Z:\HPCHEM1\BNA F\DATA\BF011615\
 Data File : BF076919.D
 Acq On : 16 Jan 2015 21:22
 Operator : IZ / TP
 Sample : G1080-01
 Misc : W
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TE-PMW-34

Quant Time: Jan 16 23:44:52 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF011415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 16 21:57:04 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.97	152	30781	20.00	ng	-0.01
21) Naphthalene-d8	10.88	136	133669	20.00	ng	0.00
38) Acenaphthene-d10	15.01	164	73389	20.00	ng	-0.01
63) Phenanthrene-d10	17.77	188	123119	20.00	ng	0.00
75) Chrysene-d12	21.27	240	106320	20.00	ng	0.00
86) Perylene-d12	22.97	264	91647	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.03	112	88929	47.50	ng	-0.01
7) Phenol-d6	7.37	99	66563	28.18	ng	0.00
23) Nitrobenzene-d5	9.27	82	201504	83.52	ng	-0.01
41) 2,4,6-Tribromophenol	16.68	330	84418	141.71	ng	-0.01
44) 2-Fluorobiphenyl	13.53	172	429274	89.02	ng	0.00
78) Terphenyl-d14	20.00	244	370825	80.30	ng	0.00
Target Compounds						Qvalue
49) Dimethylphthalate	14.60	163	17197	3.30	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF011615\
Data File : BF076919.D
Acq On : 16 Jan 2015 21:22
Operator : IZ / TP
Sample : G1080-01
Misc : W
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TE-PMW-34

Quant Time: Jan 16 23:44:52 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF011415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 16 21:57:04 2015
Response via : Initial Calibration

