

Data Path : U:\HPCHEM1\BNA F\DATA\BF011618\
 Data File : BF102099.D
 Acq On : 16 Jan 2018 10:40
 Operator : SJ/JU
 Sample : J1116-03
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-41-4.5-5.0

Quant Time: Jan 16 23:44:49 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF010918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 11 12:53:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.72	152	158483	20.00	ng	0.00
21) Naphthalene-d8	8.00	136	663273	20.00	ng	0.00
38) Acenaphthene-d10	9.76	164	303526	20.00	ng	0.00
63) Phenanthrene-d10	11.23	188	522138	20.00	ng	-0.01
75) Chrysene-d12	13.87	240	362140	20.00	ng	-0.01
86) Perylene-d12	15.28	264	284367	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.34	112	1177325	104.90	ng	0.01
7) Phenol-d6	6.36	99	1421333	106.14	ng	0.00
23) Nitrobenzene-d5	7.29	82	899880	79.73	ng	-0.01
41) 2,4,6-Tribromophenol	10.55	330	317706	119.94	ng	0.00
44) 2-Fluorobiphenyl	9.08	172	1297406	66.78	ng	-0.01
78) Terphenyl-d14	12.83	244	1179541	67.63	ng	0.00
Target Compounds						
10) Phenol	6.37	94	46303	3.060	ng	99
49) Dimethylphthalate	9.47	163	238742	9.805	ng	99
83) Bis(2-ethylhexyl)phthalate	13.86	149	97392	5.779	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : U:\HPCHEM1\BNA F\DATA\BF011618\
Data File : BF102099.D
Acq On : 16 Jan 2018 10:40
Operator : SJ/JU
Sample : J1116-03
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-41-4.5-5.0

Quant Time: Jan 16 23:44:49 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF010918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 11 12:53:54 2018
Response via : Initial Calibration

