

Data Path : U:\HPCHEM1\BNA F\DATA\BF011618\  
 Data File : BF102133.D  
 Acq On : 17 Jan 2018 1:56  
 Operator : SJ/JU  
 Sample : J1122-14  
 Misc :  
 ALS Vial : 22 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 C-2

Quant Time: Jan 17 04:05:43 2018  
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF010918.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Jan 11 12:53:54 2018  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min)       |
|-----------------------------|-------|------|----------|--------|-------|----------------|
| 1) 1,4-Dichlorobenzene-d4   | 6.72  | 152  | 137685   | 20.00  | ng    | 0.00           |
| 21) Naphthalene-d8          | 8.00  | 136  | 550657   | 20.00  | ng    | -0.01          |
| 38) Acenaphthene-d10        | 9.75  | 164  | 233011   | 20.00  | ng    | -0.01          |
| 63) Phenanthrene-d10        | 11.23 | 188  | 359287   | 20.00  | ng    | -0.01          |
| 75) Chrysene-d12            | 13.86 | 240  | 249528   | 20.00  | ng    | -0.02          |
| 86) Perylene-d12            | 15.28 | 264  | 249495   | 20.00  | ng    | -0.01          |
| System Monitoring Compounds |       |      |          |        |       |                |
| 5) 2-Fluorophenol           | 5.33  | 112  | 1008981  | 103.48 | ng    | 0.00           |
| 7) Phenol-d6                | 6.36  | 99   | 1244812  | 107.00 | ng    | 0.00           |
| 23) Nitrobenzene-d5         | 7.29  | 82   | 772484   | 82.44  | ng    | -0.01          |
| 41) 2,4,6-Tribromophenol    | 10.54 | 330  | 241544   | 118.78 | ng    | -0.01          |
| 44) 2-Fluorobiphenyl        | 9.08  | 172  | 1374927  | 92.18  | ng    | -0.01          |
| 78) Terphenyl-d14           | 12.82 | 244  | 957825   | 79.70  | ng    | -0.01          |
| Target Compounds            |       |      |          |        |       |                |
| 49) Dimethylphthalate       | 9.47  | 163  | 171283   | 9.163  | ng    | Qvalue<br># 99 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : U:\HPCHEM1\BNA F\DATA\BF011618\  
Data File : BF102133.D  
Acq On : 17 Jan 2018 1:56  
Operator : SJ/JU  
Sample : J1122-14  
Misc :  
ALS Vial : 22 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
C-2

Quant Time: Jan 17 04:05:43 2018  
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF010918.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Jan 11 12:53:54 2018  
Response via : Initial Calibration

