

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF011624\
 Data File : BF137089.D
 Acq On : 16 Jan 2024 20:06
 Operator : CG\JU
 Sample : P1146-05
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 3RD-FLOOR-SAMPLE-01

Quant Time: Jan 17 02:02:46 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 21 01:54:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.781	152	50173	20.000	ng	-0.01
21) Naphthalene-d8	8.057	136	185005	20.000	ng	-0.01
39) Acenaphthene-d10	9.816	164	85007	20.000	ng	-0.01
64) Phenanthrene-d10	11.310	188	140535	20.000	ng	0.00
76) Chrysene-d12	13.974	240	111714	20.000	ng	0.00
86) Perylene-d12	15.468	264	119300	20.000	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.475	112	48051	14.943	ng	0.05
7) Phenol-d6	6.451	99	218136	52.352	ng	0.01
23) Nitrobenzene-d5	7.345	82	168662	46.651	ng	-0.01
42) 2,4,6-Tribromophenol	10.616	330	9203	9.192	ng	0.00
45) 2-Fluorobiphenyl	9.133	172	328298	51.943	ng	-0.02
79) Terphenyl-d14	12.904	244	293491	36.714	ng	-0.01
Target Compounds						Qvalue
74) Di-n-butylphthalate	11.874	149	223914	26.651	ng	100
80) Butylbenzylphthalate	13.386	149	25076	6.284	ng	97
84) Bis(2-ethylhexyl)phtha...	13.951	149	67012	13.347	ng	# 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF011624\
Data File : BF137089.D
Acq On : 16 Jan 2024 20:06
Operator : CG\JU
Sample : P1146-05
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
3RD-FLOOR-SAMPLE-01

Quant Time: Jan 17 02:02:46 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 21 01:54:25 2023
Response via : Initial Calibration

