




Data Path : Z:\HPCHEM1\BNA_F\DATA\BF011717\
Data File : BF092317.D
Acq On : 17 Jan 2017 15:11
Operator : UM/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 18 07:12:53 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF122116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 18 00:11:54 2017
Response via : Initial Calibration



TIC: BF092317.D

(69) Pentachlorophenol (C)




10.885min (-0.000) 38267.70ng

response 299808

Ion	Exp%	Act%
265.70	100	100
268.00	66.60	64.81
264.00	62.90	62.72
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF011717\
Data File : BF092317.D
Acq On : 17 Jan 2017 15:11
Operator : UM/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 18 07:12:53 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF122116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 18 00:11:54 2017
Response via : Initial Calibration



TIC: BF092317.D

(76) Benzidine

12.405min (-0.000) -20.00ng

response 1635301

Ion	Exp%	Act%
184.00	100	100
185.00	16.70	15.73
183.00	11.50	11.56
0.00	0.00	0.00