




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF011725\
Data File : BF141192.D
Acq On : 17 Jan 2025 14:59
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 17 15:59:03 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 10 22:54:19 2025
Response via : Initial Calibration



TIC: BF141192.D\data.ms

(70) Pentachlorophenol (C)




11.128min (-0.018) 63009.31 ng

response 214616

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.00	64.07
264.00	62.70	62.41
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF011725\
Data File : BF141192.D
Acq On : 17 Jan 2025 14:59
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 17 15:59:03 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 10 22:54:19 2025
Response via : Initial Calibration



TIC: BF141192.D\data.ms

(77) Benzidine**12.663min (-0.018) 84932.11 ng****response 826882**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.20	14.40
183.00	11.50	12.05
0.00	0.00	0.00