

Data Path : Z:\HPCHEM1\BNA\_F\DATA\BF011817\  
 Data File : BF092346.D  
 Acq On : 18 Jan 2017 12:35  
 Operator : UM/SJ  
 Sample : I1246-03  
 Misc :  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 ROW-SP-0-2-COMP

Quant Time: Jan 19 01:37:49 2017  
 Quant Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF122116.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Jan 19 01:06:01 2017  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|-------|------|----------|--------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 6.54  | 152  | 186078   | 20.00  | ng    | 0.00     |
| 21) Naphthalene-d8          | 7.83  | 136  | 721147   | 20.00  | ng    | 0.00     |
| 38) Acenaphthene-d10        | 9.57  | 164  | 427249   | 20.00  | ng    | -0.01    |
| 63) Phenanthrene-d10        | 11.06 | 188  | 658465   | 20.00  | ng    | 0.00     |
| 75) Chrysene-d12            | 13.69 | 240  | 400443   | 20.00  | ng    | -0.01    |
| 86) Perylene-d12            | 15.03 | 264  | 392080   | 20.00  | ng    | 0.00     |
| System Monitoring Compounds |       |      |          |        |       |          |
| 5) 2-Fluorophenol           | 5.15  | 112  | 1289198  | 115.68 | ng    | 0.02     |
| 7) Phenol-d6                | 6.21  | 99   | 1454655  | 106.91 | ng    | -0.01    |
| 23) Nitrobenzene-d5         | 7.11  | 82   | 1037218  | 79.98  | ng    | -0.01    |
| 41) 2,4,6-Tribromophenol    | 10.37 | 330  | 473927   | 134.04 | ng    | 0.00     |
| 44) 2-Fluorobiphenyl        | 8.91  | 172  | 2132093  | 105.33 | ng    | -0.01    |
| 78) Terphenyl-d14           | 12.64 | 244  | 1735018  | 105.08 | ng    | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
|------------------|--------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA\_F\DATA\BF011817\  
Data File : BF092346.D  
Acq On : 18 Jan 2017 12:35  
Operator : UM/SJ  
Sample : I1246-03  
Misc :  
ALS Vial : 6 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
ROW-SP-0-2-COMP

Quant Time: Jan 19 01:37:49 2017  
Quant Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF122116.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Jan 19 01:06:01 2017  
Response via : Initial Calibration

