

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF011821\
 Data File : BF122989.D
 Acq On : 18 Jan 2021 15:45
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 18 16:19:48 2021
 Quant Title :
 QLast Update : Mon Jan 18 13:14:08 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
2	1,4-Dioxane	0.643	0.597	7.2	99	-0.01
3	Pyridine	1.739	1.636	5.9	100	0.00
4	n-Nitrosodimethylamine	0.776	0.702	9.5	96	-0.01
5 S	2-Fluorophenol	1.323	1.224	7.5	99	0.00
6	Aniline	2.145	1.941	9.5	96	0.00
7 S	Phenol-d6	1.778	1.616	9.1	96	0.00
8	2-Chlorophenol	1.397	1.282	8.2	97	0.00
9	Benzaldehyde	1.042	0.878	15.7	97	0.00
10 C	Phenol	1.786	1.626	9.0	96	0.00
11	bis(2-Chloroethyl)ether	1.446	1.290	10.8	95	0.00
12	1,3-Dichlorobenzene	1.617	1.457	9.9	96	0.00
13 C	1,4-Dichlorobenzene	1.628	1.454	10.7	95	0.00
14	1,2-Dichlorobenzene	1.519	1.372	9.7	96	0.00
15	Benzyl Alcohol	1.253	1.164	7.1	97	0.00
16	2,2'-oxybis(1-Chloropropane	2.520	2.151	14.6	90	0.00
17	2-Methylphenol	1.185	1.064	10.2	96	0.00
18	Hexachloroethane	0.589	0.538	8.7	97	0.00
19 P	n-Nitroso-di-n-propylamine	1.053	0.913	13.3	91	0.00
20	3+4-Methylphenols	1.483	1.342	9.5	95	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	99	0.00
22	Acetophenone	0.503	0.443	11.9	95	0.00
23 S	Nitrobenzene-d5	0.389	0.350	10.0	95	0.00
24	Nitrobenzene	0.398	0.361	9.3	96	0.00
25	Isophorone	0.693	0.594	14.3	92	0.00
26 C	2-Nitrophenol	0.180	0.171	5.0	96	0.00
27	2,4-Dimethylphenol	0.287	0.255	11.1	94	0.00
28	bis(2-Chloroethoxy)methane	0.469	0.403	14.1	91	0.00
29 C	2,4-Dichlorophenol	0.299	0.266	11.0	93	0.00
30	1,2,4-Trichlorobenzene	0.319	0.283	11.3	95	0.00
31	Naphthalene	1.028	0.904	12.1	94	0.00
32	Benzoic acid	0.241	0.224	7.1	94	-0.01
33	4-Chloroaniline	0.449	0.387	13.8	93	0.00
34 C	Hexachlorobutadiene	0.174	0.155	10.9	94	0.00
35	Caprolactam	0.102	0.088	13.7	90	-0.01
36 C	4-Chloro-3-methylphenol	0.318	0.279	12.3	93	0.00
37	2-Methylnaphthalene	0.709	0.610	14.0	90	0.00
38	1-Methylnaphthalene	0.670	0.576	14.0	92	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	95	0.00
40	1,2,4,5-Tetrachlorobenzene	0.549	0.482	12.2	92	0.00
41 P	Hexachlorocyclopentadiene	0.267	0.236	11.6	87	0.00
42 S	2,4,6-Tribromophenol	0.226	0.203	10.2	90	0.00
43 C	2,4,6-Trichlorophenol	0.409	0.371	9.3	91	0.00
44	2,4,5-Trichlorophenol	0.415	0.381	8.2	94	0.00
45 S	2-Fluorobiphenyl	1.428	1.190	16.7	91	0.00
46	1,1'-Biphenyl	1.627	1.425	12.4	92	0.00
47	2-Chloronaphthalene	1.357	1.202	11.4	90	0.00
48	2-Nitroaniline	0.463	0.423	8.6	92	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF011821\
 Data File : BF122989.D
 Acq On : 18 Jan 2021 15:45
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 18 16:19:48 2021
 Quant Title :
 QLast Update : Mon Jan 18 13:14:08 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49	Acenaphthylene	1.980	1.754	11.4	92	0.00
50	Dimethylphthalate	1.552	1.309	15.7	89	0.00
51	2,6-Dinitrotoluene	0.340	0.299	12.1	88	0.00
52 C	Acenaphthene	1.280	1.129	11.8	91	0.00
53	3-Nitroaniline	0.420	0.376	10.5	91	0.00
54 P	2,4-Dinitrophenol	0.129	0.128	0.8	87	0.00
55	Dibenzofuran	1.784	1.552	13.0	90	0.00
56 P	4-Nitrophenol	0.309	0.292	5.5	93	0.00
57	2,4-Dinitrotoluene	0.437	0.398	8.9	92	0.00
58	Fluorene	1.395	1.191	14.6	89	0.00
59	2,3,4,6-Tetrachlorophenol	0.331	0.299	9.7	92	0.00
60	Diethylphthalate	1.509	1.279	15.2	88	0.00
61	4-Chlorophenyl-phenylether	0.627	0.546	12.9	91	0.00
62	4-Nitroaniline	0.414	0.372	10.1	92	0.00
63	Azobenzene	1.510	1.311	13.2	90	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	96	0.00
65	4,6-Dinitro-2-methylphenol	0.113	0.097	14.2	88	0.00
66 c	n-Nitrosodiphenylamine	0.747	0.638	14.6	91	0.00
67	4-Bromophenyl-phenylether	0.245	0.208	15.1	88	0.00
68	Hexachlorobenzene	0.280	0.239	14.6	90	0.00
69	Atrazine	0.199	0.172	13.6	89	0.00
70 C	Pentachlorophenol	0.147	0.138	6.1	91	0.00
71	Phenanthrene	1.199	1.029	14.2	91	0.00
72	Anthracene	1.185	1.017	14.2	90	0.00
73	Carbazole	1.148	0.994	13.4	92	0.00
74	Di-n-butylphthalate	1.359	1.141	16.0	86	0.00
75 C	Fluoranthene	1.213	1.038	14.4	90	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	94	0.00
77	Benzydine	0.530	0.539	-1.7	97	0.00
78	Pyrene	1.601	1.424	11.1	90	0.00
79 S	Terphenyl-d14	1.091	0.943	13.6	90	0.00
80	Butylbenzylphthalate	0.718	0.644	10.3	87	0.00
81	Benzo(a)anthracene	1.339	1.182	11.7	91	0.00
82	3,3'-Dichlorobenzidine	0.438	0.407	7.1	90	0.00
83	Chrysene	1.350	1.163	13.9	88	0.00
84	Bis(2-ethylhexyl)phthalate	0.957	0.797	16.7	83	0.00
85 c	Di-n-octyl phthalate	1.548	1.328	14.2	82	0.00
86 I	Perylene-d12	1.000	1.000	0.0	87	0.00
87	Indeno(1,2,3-cd)pyrene	1.260	1.165	7.5	82	-0.01
88	Benzo(b)fluoranthene	1.350	1.344	0.4	91	0.00
89	Benzo(k)fluoranthene	1.384	1.158	16.3	77	0.00
90 C	Benzo(a)pyrene	1.219	1.137	6.7	83	0.00
91	Dibenzo(a,h)anthracene	1.032	0.957	7.3	81	0.00
92	Benzo(g,h,i)perylene	0.999	0.929	7.0	85	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_F

LabSampleId :

SSTDCCC040