

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF011821\
 Data File : BF122989.D
 Acq On : 18 Jan 2021 15:45
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 18 16:19:48 2021
 Quant Title :
 QLast Update : Mon Jan 18 13:14:08 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	98	0.00
2	1,4-Dioxane	40.000	37.172	7.1	99	-0.01
3	Pyridine	40.000	37.616	6.0	100	0.00
4	n-Nitrosodimethylamine	40.000	36.155	9.6	96	-0.01
5 S	2-Fluorophenol	80.000	74.018	7.5	99	0.00
6	Aniline	40.000	36.192	9.5	96	0.00
7 S	Phenol-d6	80.000	72.702	9.1	96	0.00
8	2-Chlorophenol	40.000	36.712	8.2	97	0.00
9	Benzaldehyde	40.000	39.900	0.3	97	0.00
10 C	Phenol	40.000	36.428	8.9	96	0.00
11	bis(2-Chloroethyl)ether	40.000	35.682	10.8	95	0.00
12	1,3-Dichlorobenzene	40.000	36.042	9.9	96	0.00
13 C	1,4-Dichlorobenzene	40.000	35.712	10.7	95	0.00
14	1,2-Dichlorobenzene	40.000	36.135	9.7	96	0.00
15	Benzyl Alcohol	40.000	37.146	7.1	97	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	34.135	14.7	90	0.00
17	2-Methylphenol	40.000	35.910	10.2	96	0.00
18	Hexachloroethane	40.000	36.594	8.5	97	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	34.688	13.3	91	0.00
20	3+4-Methylphenols	40.000	36.204	9.5	95	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	99	0.00
22	Acetophenone	40.000	35.244	11.9	95	0.00
23 S	Nitrobenzene-d5	80.000	71.993	10.0	95	0.00
24	Nitrobenzene	40.000	36.292	9.3	96	0.00
25	Isophorone	40.000	34.237	14.4	92	0.00
26 C	2-Nitrophenol	40.000	38.054	4.9	96	0.00
27	2,4-Dimethylphenol	40.000	35.666	10.8	94	0.00
28	bis(2-Chloroethoxy)methane	40.000	34.366	14.1	91	0.00
29 C	2,4-Dichlorophenol	40.000	35.519	11.2	93	0.00
30	1,2,4-Trichlorobenzene	40.000	35.491	11.3	95	0.00
31	Naphthalene	40.000	35.172	12.1	94	0.00
32	Benzoic acid	40.000	37.252	6.9	94	-0.01
33	4-Chloroaniline	40.000	34.547	13.6	93	0.00
34 C	Hexachlorobutadiene	40.000	35.537	11.2	94	0.00
35	Caprolactam	40.000	34.775	13.1	90	-0.01
36 C	4-Chloro-3-methylphenol	40.000	35.105	12.2	93	0.00
37	2-Methylnaphthalene	40.000	34.390	14.0	90	0.00
38	1-Methylnaphthalene	40.000	34.406	14.0	92	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	95	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	35.143	12.1	92	0.00
41 P	Hexachlorocyclopentadiene	40.000	35.338	11.7	87	0.00
42 S	2,4,6-Tribromophenol	80.000	71.712	10.4	90	0.00
43 C	2,4,6-Trichlorophenol	40.000	36.212	9.5	91	0.00
44	2,4,5-Trichlorophenol	40.000	36.671	8.3	94	0.00
45 S	2-Fluorobiphenyl	80.000	66.639	16.7	91	0.00
46	1,1'-Biphenyl	40.000	35.041	12.4	92	0.00
47	2-Chloronaphthalene	40.000	35.420	11.4	90	0.00
48	2-Nitroaniline	40.000	36.579	8.6	92	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF011821\
 Data File : BF122989.D
 Acq On : 18 Jan 2021 15:45
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 18 16:19:48 2021
 Quant Title :
 QLast Update : Mon Jan 18 13:14:08 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49	Acenaphthylene	40.000	35.433	11.4	92	0.00
50	Dimethylphthalate	40.000	33.737	15.7	89	0.00
51	2,6-Dinitrotoluene	40.000	35.208	12.0	88	0.00
52 C	Acenaphthene	40.000	35.278	11.8	91	0.00
53	3-Nitroaniline	40.000	35.856	10.4	91	0.00
54 P	2,4-Dinitrophenol	40.000	36.269	9.3	87	0.00
55	Dibenzofuran	40.000	34.810	13.0	90	0.00
56 P	4-Nitrophenol	40.000	37.793	5.5	93	0.00
57	2,4-Dinitrotoluene	40.000	36.408	9.0	92	0.00
58	Fluorene	40.000	34.150	14.6	89	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	36.117	9.7	92	0.00
60	Diethylphthalate	40.000	33.911	15.2	88	0.00
61	4-Chlorophenyl-phenylether	40.000	34.798	13.0	91	0.00
62	4-Nitroaniline	40.000	35.963	10.1	92	0.00
63	Azobenzene	40.000	34.744	13.1	90	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	96	0.00
65	4,6-Dinitro-2-methylphenol	40.000	34.345	14.1	88	0.00
66 c	n-Nitrosodiphenylamine	40.000	34.206	14.5	91	0.00
67	4-Bromophenyl-phenylether	40.000	34.061	14.8	88	0.00
68	Hexachlorobenzene	40.000	34.162	14.6	90	0.00
69	Atrazine	40.000	34.627	13.4	89	0.00
70 C	Pentachlorophenol	40.000	37.423	6.4	91	0.00
71	Phenanthrene	40.000	34.326	14.2	91	0.00
72	Anthracene	40.000	34.335	14.2	90	0.00
73	Carbazole	40.000	34.637	13.4	92	0.00
74	Di-n-butylphthalate	40.000	33.586	16.0	86	0.00
75 C	Fluoranthene	40.000	34.254	14.4	90	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	94	0.00
77	Benzydine	40.000	40.699	-1.7	97	0.00
78	Pyrene	40.000	35.565	11.1	90	0.00
79 S	Terphenyl-d14	80.000	69.127	13.6	90	0.00
80	Butylbenzylphthalate	40.000	35.900	10.3	87	0.00
81	Benzo(a)anthracene	40.000	35.320	11.7	91	0.00
82	3,3'-Dichlorobenzidine	40.000	37.157	7.1	90	0.00
83	Chrysene	40.000	34.456	13.9	88	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	33.331	16.7	83	0.00
85 c	Di-n-octyl phthalate	40.000	34.316	14.2	82	0.00
86 I	Perylene-d12	20.000	20.000	0.0	87	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	36.985	7.5	82	-0.01
88	Benzo(b)fluoranthene	40.000	39.824	0.4	91	0.00
89	Benzo(k)fluoranthene	40.000	33.472	16.3	77	0.00
90 C	Benzo(a)pyrene	40.000	37.299	6.8	83	0.00
91	Dibenzo(a,h)anthracene	40.000	37.083	7.3	81	0.00
92	Benzo(g,h,i)perylene	40.000	37.173	7.1	85	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_F

LabSampleId :

SSTDCCC040