

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF011822\
 Data File : BF126603.D
 Acq On : 18 Jan 2022 14:55
 Operator : CG/JU
 Sample : PB141938BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB141938BL

Quant Time: Jan 19 07:06:12 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 19 07:04:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.963	152	118134	20.000	ng	0.00
21) Naphthalene-d8	8.245	136	485022	20.000	ng	0.00
39) Acenaphthene-d10	10.004	164	288197	20.000	ng	0.00
64) Phenanthrene-d10	11.498	188	525676	20.000	ng	0.00
76) Chrysene-d12	14.174	240	454291	20.000	ng	# 0.02
86) Perylene-d12	15.745	264	407166	20.000	ng	0.07
System Monitoring Compounds						
5) 2-Fluorophenol	5.593	112	1038641	115.695	ng	0.01
7) Phenol-d6	6.575	99	1429393	114.598	ng	0.00
23) Nitrobenzene-d5	7.522	82	1033664	83.416	ng	-0.01
42) 2,4,6-Tribromophenol	10.798	330	337359	125.083	ng	0.00
45) 2-Fluorobiphenyl	9.322	172	1421574	75.543	ng	0.00
79) Terphenyl-d14	13.098	244	1839302	67.908	ng	0.00
Target Compounds						Qvalue
3) Pyridine	3.769	79	1354	0.109	ng	# 73
6) Aniline	6.628	93	133	0.009	ng	# 42
9) Benzaldehyde	6.581	77	2861	Below Cal		# 3
10) Phenol	6.575	94	116	0.009	ng	# 1
11) bis(2-Chloroethyl)ether	6.628	93	133	0.012	ng	# 19
15) Benzyl Alcohol	6.963	79	4812	0.433	ng	# 26
19) n-Nitroso-di-n-propyla...	7.522	70	143468	15.760	ng	# 70
24) Nitrobenzene	7.522	77	3058	0.239	ng	# 36
25) Isophorone	7.522	82	1033780	43.018	ng	# 69
38) 1-Methylnaphthalene	9.322	142	531	0.034	ng	# 84
46) 1,1'-Biphenyl	9.322	154	2527	0.116	ng	# 1
48) 2-Nitroaniline	9.322	65	3618	0.614	ng	# 1
51) 2,6-Dinitrotoluene	10.004	165	36566	10.352	ng	# 21
52) Acenaphthene	10.004	154	830	0.051	ng	# 6
57) 2,4-Dinitrotoluene	10.004	165	36566	8.543	ng	# 15
58) Fluorene	10.798	166	13520	0.728	ng	# 88
63) Azobenzene	10.792	77	2988	0.099	ng	# 27
65) 4,6-Dinitro-2-methylph...	10.792	198	716	7.613	ng	# 1
66) n-Nitrosodiphenylamine	10.798	169	11506	0.648	ng	# 52
67) 4-Bromophenyl-phenylether	10.798	248	23304	3.847	ng	# 1
70) Pentachlorophenol	11.251	266	164	0.044	ng	# 1
74) Di-n-butylphthalate	12.057	149	242	0.007	ng	# 79
77) Benzidine	13.098	184	25503	1.627	ng	# 97
78) Pyrene	13.098	202	5931	0.161	ng	# 72
81) Benzo(a)anthracene	14.174	228	1247	0.040	ng	# 59
83) Chrysene	14.174	228	1247	0.042	ng	# 57
84) Bis(2-ethylhexyl)phtha...	14.151	149	3095	0.143	ng	# 88
85) Di-n-octyl phthalate	14.845	149	234	0.007	ng	# 59
90) Benzo(a)pyrene	15.751	252	1302	0.059	ng	# 64


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF011822\
Data File : BF126603.D
Acq On : 18 Jan 2022 14:55
Operator : CG/JU
Sample : PB141938BL
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB141938BL

Quant Time: Jan 19 07:06:12 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 19 07:04:14 2022
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.963 min Scan# 81
Delta R.T. -0.006 min
Lab File: BF126603.D
Acq: 18 Jan 2022 14:55




Instrument :
BNA_F
ClientSampleId :
PB141938BL


Tgt Ion:152 Resp: 118134
Ion Ratio Lower Upper
152 100
150 160.7 128.1 192.1
115 73.5 56.7 85.1



#3
Pyridine
Concen: 0.109 ng
RT: 3.769 min Scan# 286
Delta R.T. 0.182 min
Lab File: BF126603.D
Acq: 18 Jan 2022 14:55

Tgt Ion: 79 Resp: 1354
Ion Ratio Lower Upper
79 100
52 61.1 56.2 84.4
51 0.0 26.7 40.1#









#5
2-Fluorophenol
Concen: 115.695 ng
RT: 5.593 min Scan# 594
Delta R.T. 0.012 min
Lab File: BF126603.D
Acq: 18 Jan 2022 14:55




Instrument :
BNA_F
ClientSampleId :
PB141938BL


Tgt Ion:112 Resp: 1038641
Ion Ratio Lower Upper
112 100
64 65.4 49.5 74.3
63 33.9 23.9 35.9



#6
Aniline
Concen: 0.009 ng
RT: 6.628 min Scan# 772
Delta R.T. -0.006 min
Lab File: BF126603.D
Acq: 18 Jan 2022 14:55

Tgt Ion: 93 Resp: 133
Ion Ratio Lower Upper
93 100
66 0.0 32.2 48.4#
65 0.0 16.3 24.5#









#7
Phenol-d6
Concen: 114.598 ng
RT: 6.575 min Scan# 70
Delta R.T. -0.006 min
Lab File: BF126603.D
Acq: 18 Jan 2022 14:55




Instrument :
BNA_F
ClientSampleId :
PB141938BL


Tgt Ion: 99 Resp: 1429393
Ion Ratio Lower Upper
99 100
42 16.7 14.2 21.4
71 41.7 24.3 36.5#



#9
Benzaldehyde
Concen: Below Cal
RT: 6.581 min Scan# 764
Delta R.T. 0.059 min
Lab File: BF126603.D
Acq: 18 Jan 2022 14:55

Tgt Ion: 77 Resp: 2861
Ion Ratio Lower Upper
77 100
105 0.0 74.9 114.9#
106 0.0 72.4 112.4#









#10
Phenol
Concen: 0.009 ng
RT: 6.575 min Scan# 701
Delta R.T. -0.018 min
Lab File: BF126603.D
Acq: 18 Jan 2022 14:55




Instrument :
BNA_F
ClientSampleId :
PB141938BL


Tgt Ion: 94 Resp: 116
Ion Ratio Lower Upper
94 100
65 3617.6 7.2 47.2#
66 18426.7 14.5 54.5#



#11
bis(2-Chloroethyl)ether
Concen: 0.012 ng
RT: 6.628 min Scan# 772
Delta R.T. -0.071 min
Lab File: BF126603.D
Acq: 18 Jan 2022 14:55

Tgt Ion: 93 Resp: 133
Ion Ratio Lower Upper
93 100
63 0.0 60.5 100.5#
95 0.0 12.9 52.9#









#15
Benzyl Alcohol
Concen: 0.433 ng
RT: 6.963 min Scan# 81
Delta R.T. -0.141 min
Lab File: BF126603.D
Acq: 18 Jan 2022 14:55




Instrument :
BNA_F
ClientSampleId :
PB141938BL


Tgt Ion: 79 Resp: 4812
Ion Ratio Lower Upper
79 100
108 12.2 63.2 94.8#
77 116.2 49.4 74.2#



#19
n-Nitroso-di-n-propylamine
Concen: 15.760 ng
RT: 7.522 min Scan# 924
Delta R.T. 0.141 min
Lab File: BF126603.D
Acq: 18 Jan 2022 14:55


Tgt Ion: 70 Resp: 143468
Ion Ratio Lower Upper
70 100
42 36.6 45.8 68.8#
101 0.0 6.9 10.3#
130 1.4 14.4 21.6#








#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.245 min Scan# 1048
Delta R.T. -0.006 min
Lab File: BF126603.D
Acq: 18 Jan 2022 14:55

Instrument :
BNA_F
ClientSampleId :
PB141938BL







Tgt Ion:136 Resp: 485022
Ion Ratio Lower Upper
136 100
137 10.7 8.6 13.0
54 8.7 7.8 11.6
68 7.2 6.3 9.5



#23
Nitrobenzene-d5
Concen: 83.416 ng
RT: 7.522 min Scan# 924
Delta R.T. -0.012 min
Lab File: BF126603.D
Acq: 18 Jan 2022 14:55


Tgt Ion: 82 Resp: 1033664
Ion Ratio Lower Upper
82 100
128 29.7 32.7 49.1#
54 41.7 41.7 62.5








#24
Nitrobenzene
Concen: 0.239 ng
RT: 7.522 min Scan# 91
Delta R.T. -0.029 min
Lab File: BF126603.D
Acq: 18 Jan 2022 14:55


Instrument :
BNA_F
ClientSampleId :
PB141938BL





Tgt Ion: 77 Resp: 3058
Ion Ratio Lower Upper
77 100
123 0.0 34.6 52.0#
65 34.2 10.6 15.8#




#25
Isophorone
Concen: 43.018 ng
RT: 7.522 min Scan# 924
Delta R.T. -0.265 min
Lab File: BF126603.D
Acq: 18 Jan 2022 14:55



Tgt Ion: 82 Resp: 1033780
Ion Ratio Lower Upper
82 100
95 0.0 5.0 7.4#
138 0.0 12.3 18.5#









#38
1-Methylnaphthalene
Concen: 0.034 ng
RT: 9.322 min Scan# 1186
Delta R.T. 0.259 min
Lab File: BF126603.D
Acq: 18 Jan 2022 14:55




Instrument :
BNA_F
ClientSampleId :
PB141938BL


Tgt Ion:142 Resp: 531
Ion Ratio Lower Upper
142 100
141 104.8 72.1 108.1



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 10.004 min Scan# 1346
Delta R.T. -0.006 min
Lab File: BF126603.D
Acq: 18 Jan 2022 14:55

Tgt Ion:164 Resp: 288197
Ion Ratio Lower Upper
164 100
162 99.7 81.4 122.0
160 44.5 36.6 54.8





#42

2,4,6-Tribromophenol

Concen: 125.083 ng

RT: 10.798 min Scan# 1481

Delta R.T. 0.000 min

Lab File: BF126603.D


Acq: 18 Jan 2022 14:55

Instrument :

BNA_F

ClientSampleId :

PB141938BL






Tgt Ion:330 Resp: 337359

Ion Ratio Lower Upper

330 100

332 97.8 76.0 114.0

141 36.8 30.1 45.1



#45

2-Fluorobiphenyl

Concen: 75.543 ng

RT: 9.322 min Scan# 1230

Delta R.T. -0.006 min

Lab File: BF126603.D

Acq: 18 Jan 2022 14:55




Tgt Ion:172 Resp: 1421574


Ion Ratio Lower Upper

172 100

171 35.3 28.3 42.5

170 23.4 19.0 28.6





#46

1,1'-Biphenyl

Concen: 0.116 ng

RT: 9.322 min Scan# 11

Delta R.T. -0.106 min

Lab File: BF126603.D

Acq: 18 Jan 2022 14:55

Instrument :

BNA_F

ClientSampleId :

PB141938BL





Tgt Ion:154 Resp: 2527

Ion Ratio Lower Upper

154 100

153 1085.8 19.0 59.0#

76 3064.8 0.0 34.5#



#48

2-Nitroaniline

Concen: 0.614 ng

RT: 9.322 min Scan# 1230

Delta R.T. -0.224 min

Lab File: BF126603.D

Acq: 18 Jan 2022 14:55




Tgt Ion: 65 Resp: 3618


Ion Ratio Lower Upper

65 100

92 195.8 54.4 81.6#

138 30.0 72.5 108.7#





#51

2,6-Dinitrotoluene

Concen: 10.352 ng

RT: 10.004 min Scan# 1346

Delta R.T. 0.212 min

Lab File: BF126603.D

Acq: 18 Jan 2022 14:55

Instrument :

BNA_F

ClientSampleId :

PB141938BL





Tgt Ion:165 Resp: 36566

Ion Ratio Lower Upper

165 100

63 1.0 48.4 72.6#

89 1.5 50.5 75.7#



#52

Acenaphthene

Concen: 0.051 ng

RT: 10.004 min Scan# 1346

Delta R.T. -0.041 min

Lab File: BF126603.D

Acq: 18 Jan 2022 14:55




Tgt Ion:154 Resp: 830


Ion Ratio Lower Upper

154 100

153 0.0 87.8 131.6#

152 0.0 39.4 59.2#









#57
2,4-Dinitrotoluene
Concen: 8.543 ng
RT: 10.004 min Scan# 1481
Delta R.T. -0.188 min
Lab File: BF126603.D
Acq: 18 Jan 2022 14:55




Instrument :
BNA_F
ClientSampleId :
PB141938BL


Tgt Ion:165 Resp: 36566
Ion Ratio Lower Upper
165 100
63 1.0 43.0 64.6#
89 1.5 71.0 106.4#
182 0.0 1.6 2.4#



#58
Fluorene
Concen: 0.728 ng
RT: 10.798 min Scan# 1481
Delta R.T. 0.235 min
Lab File: BF126603.D
Acq: 18 Jan 2022 14:55

Tgt Ion:166 Resp: 13520
Ion Ratio Lower Upper
166 100
165 104.0 77.7 116.5
167 33.1 10.7 16.1#









#63
Azobenzene
Concen: 0.099 ng
RT: 10.792 min Scan# 1466
Delta R.T. 0.082 min
Lab File: BF126603.D
Acq: 18 Jan 2022 14:55




Instrument :
BNA_F
ClientSampleId :
PB141938BL


Tgt Ion: 77	Resp: 2988
Ion Ratio	Lower Upper
77 100	
182 0.0	0.0 39.0
105 72.2	3.2 43.2#
51 65.7	8.2 48.2#



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.498 min Scan# 1600
Delta R.T. 0.000 min
Lab File: BF126603.D
Acq: 18 Jan 2022 14:55

Tgt Ion: 188	Resp: 525676
Ion Ratio	Lower Upper
188 100	
94 12.5	10.1 15.1
80 14.3	11.8 17.6









#65
4,6-Dinitro-2-methylphenol
Concen: 7.613 ng
RT: 10.792 min Scan# 1447
Delta R.T. 0.194 min
Lab File: BF126603.D
Acq: 18 Jan 2022 14:55




Instrument :
BNA_F
ClientSampleId :
PB141938BL


Tgt Ion:198 Resp: 716
Ion Ratio Lower Upper
198 100
51 173.5 32.0 72.0#
105 190.5 32.7 72.7#



#66
n-Nitrosodiphenylamine
Concen: 0.648 ng
RT: 10.798 min Scan# 1481
Delta R.T. 0.129 min
Lab File: BF126603.D
Acq: 18 Jan 2022 14:55

Tgt Ion:169 Resp: 11506
Ion Ratio Lower Upper
169 100
168 8.1 53.4 80.0#
167 37.7 29.4 44.0









#67
4-Bromophenyl-phenylether
Concen: 3.847 ng
RT: 10.798 min Scan# 141
Delta R.T. -0.247 min
Lab File: BF126603.D
Acq: 18 Jan 2022 14:55




Instrument :
BNA_F
ClientSampleId :
PB141938BL


Tgt Ion:248 Resp: 23304
Ion Ratio Lower Upper
248 100
250 203.7 76.2 114.4#
141 475.4 66.2 99.4#



#70
Pentachlorophenol
Concen: 0.044 ng
RT: 11.251 min Scan# 1558
Delta R.T. -0.041 min
Lab File: BF126603.D
Acq: 18 Jan 2022 14:55

Tgt Ion:266 Resp: 164
Ion Ratio Lower Upper
266 100
268 226.5 52.3 78.5#
264 0.0 50.4 75.6#









#74
Di-n-butylphthalate
Concen: 0.007 ng
RT: 12.057 min Scan# 10
Delta R.T. 0.000 min
Lab File: BF126603.D
Acq: 18 Jan 2022 14:55




Instrument :
BNA_F
ClientSampleId :
PB141938BL


Tgt Ion:149 Resp: 242
Ion Ratio Lower Upper
149 100
150 0.0 7.2 10.8#
104 0.0 4.2 6.2#



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.174 min Scan# 2055
Delta R.T. 0.024 min
Lab File: BF126603.D
Acq: 18 Jan 2022 14:55

Tgt Ion:240 Resp: 454291
Ion Ratio Lower Upper
240 100
120 10.8 14.4 21.6#
236 26.3 20.5 30.7









#77
Benzidine
Concen: 1.627 ng
RT: 13.098 min Scan# 1872
Delta R.T. 0.265 min
Lab File: BF126603.D
Acq: 18 Jan 2022 14:55




Instrument :
BNA_F
ClientSampleId :
PB141938BL


Tgt Ion:184 Resp: 25503
Ion Ratio Lower Upper
184 100
185 13.7 11.8 17.8
183 8.9 8.2 12.2



#78
Pyrene
Concen: 0.161 ng
RT: 13.098 min Scan# 1872
Delta R.T. 0.147 min
Lab File: BF126603.D
Acq: 18 Jan 2022 14:55

Tgt Ion:202 Resp: 5931
Ion Ratio Lower Upper
202 100
200 42.3 16.6 24.8#
203 14.3 13.8 20.8









#79
 Terphenyl-d14
 Concen: 67.908 ng
 RT: 13.098 min Scan# 11
 Delta R.T. 0.006 min
 Lab File: BF126603.D
 Acq: 18 Jan 2022 14:55




Instrument :
 BNA_F
 ClientSampleId :
 PB141938BL


Tgt Ion:244 Resp: 1839302
 Ion Ratio Lower Upper
 244 100
 212 7.2 5.0 7.4
 122 9.2 13.3 19.9#



#81
 Benzo(a)anthracene
 Concen: 0.040 ng
 RT: 14.174 min Scan# 2055
 Delta R.T. 0.035 min
 Lab File: BF126603.D
 Acq: 18 Jan 2022 14:55

Tgt Ion:228 Resp: 1247
 Ion Ratio Lower Upper
 228 100
 226 0.0 21.4 32.0#
 229 30.9 15.4 23.2#





#83

Chrysene

Concen: 0.042 ng

RT: 14.174 min Scan# 2055

Delta R.T. -0.006 min

Lab File: BF126603.D

Acq: 18 Jan 2022 14:55

Instrument :

BNA_F

ClientSampleId :

PB141938BL





Tgt Ion:228 Resp: 1247

Ion Ratio Lower Upper

228 100

226 0.0 23.7 35.5#

229 30.9 15.5 23.3#



#84

Bis(2-ethylhexyl)phthalate

Concen: 0.143 ng

RT: 14.151 min Scan# 2051

Delta R.T. 0.029 min

Lab File: BF126603.D

Acq: 18 Jan 2022 14:55




Tgt Ion:149 Resp: 3095


Ion Ratio Lower Upper

149 100

167 20.4 21.4 32.2#


279 0.0 2.4 3.6#








#85
Di-n-octyl phthalate
Concen: 0.007 ng
RT: 14.845 min Scan# 2153
Delta R.T. 0.094 min
Lab File: BF126603.D
Acq: 18 Jan 2022 14:55


Instrument :
BNA_F
ClientSampleId :
PB141938BL





Tgt Ion:149 Resp: 234
Ion Ratio Lower Upper
149 100
167 50.9 1.1 1.7#
43 0.0 9.9 14.9#




#86
Perylene-d12
Concen: 20.000 ng
RT: 15.745 min Scan# 2322
Delta R.T. 0.071 min
Lab File: BF126603.D
Acq: 18 Jan 2022 14:55




Tgt Ion:264 Resp: 407166
Ion Ratio Lower Upper
264 100
260 24.5 19.2 28.8
265 21.3 17.4 26.2





#90
Benzo(a)pyrene
Concen: 0.059 ng
RT: 15.751 min Scan# 21
Delta R.T. 0.135 min
Lab File: BF126603.D
Acq: 18 Jan 2022 14:55

Instrument :
BNA_F
ClientSampleId :
PB141938BL



Tgt Ion:252 Resp: 1302
Ion Ratio Lower Upper
252 100
253 37.6 17.1 25.7#
125 0.0 13.0 19.6#

