


Benzaldehyde

Response Ratio




$$\text{Response} = 1.135\text{e}+000 * \text{Amt} + 3.582\text{e}-001$$

Coef of Det (r^2) = 0.991115 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF011822.M


Calibration Table Last Updated: Wed Jan 19 07:04:14 2022

2,4-Dinitrophenol



$R = 1.209e-002 A^2 + 7.370e-002 A - 1.468e-002$
 Coef of Det (r^2) = 0.998085 Curve Fit: Quadratic
 Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF011822.M
 Calibration Table Last Updated: Wed Jan 19 07:04:14 2022

4,6-Dinitro-2-methylphenol



Response = $1.129 \times 10^{-1} \times \text{Amt} - 4.161 \times 10^{-2}$
Coef of Det (r^2) = 0.992677 Curve Fit: Linear
Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF011822.M
Calibration Table Last Updated: Wed Jan 19 07:04:14 2022