

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF011822\
 Data File : BF126602.D
 Acq On : 18 Jan 2022 13:37
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF011822

Quant Time: Jan 19 07:05:54 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 19 07:04:14 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	103	0.00
2	1,4-Dioxane	40.000	35.488	11.3	98	0.00
3	Pyridine	40.000	35.531	11.2	99	0.00
4	n-Nitrosodimethylamine	40.000	35.233	11.9	99	0.00
5 S	2-Fluorophenol	80.000	70.415	12.0	100	0.00
6	Aniline	40.000	35.578	11.1	100	0.00
7 S	Phenol-d6	80.000	70.708	11.6	100	0.00
8	2-Chlorophenol	40.000	35.436	11.4	100	0.00
9	Benzaldehyde	40.000	38.491	3.8	102	0.00
10 C	Phenol	40.000	35.475	11.3	100	0.00
11	bis(2-Chloroethyl)ether	40.000	36.079	9.8	102	0.00
12	1,3-Dichlorobenzene	40.000	35.763	10.6	100	0.00
13 C	1,4-Dichlorobenzene	40.000	36.179	9.6	102	0.00
14	1,2-Dichlorobenzene	40.000	35.573	11.1	101	0.00
15	Benzyl Alcohol	40.000	35.806	10.5	100	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	35.092	12.3	98	0.00
17	2-Methylphenol	40.000	35.880	10.3	101	0.00
18	Hexachloroethane	40.000	36.443	8.9	102	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	34.572	13.6	99	0.00
20	3+4-Methylphenols	40.000	35.303	11.7	100	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	102	0.00
22	Acetophenone	40.000	34.423	13.9	101	0.00
23 S	Nitrobenzene-d5	80.000	76.296	4.6	105	0.00
24	Nitrobenzene	40.000	37.450	6.4	103	0.00
25	Isophorone	40.000	34.485	13.8	100	0.00
26 C	2-Nitrophenol	40.000	42.552	-6.4	111	0.00
27	2,4-Dimethylphenol	40.000	34.288	14.3	100	0.00
28	bis(2-Chloroethoxy)methane	40.000	34.546	13.6	101	0.00
29 C	2,4-Dichlorophenol	40.000	34.876	12.8	100	0.00
30	1,2,4-Trichlorobenzene	40.000	35.856	10.4	103	0.00
31	Naphthalene	40.000	34.686	13.3	100	0.00
32	Benzoic acid	40.000	36.103	9.7	109	0.00
33	4-Chloroaniline	40.000	34.185	14.5	99	0.00
34 C	Hexachlorobutadiene	40.000	35.430	11.4	101	0.00
35	Caprolactam	40.000	35.503	11.2	101	0.00
36 C	4-Chloro-3-methylphenol	40.000	35.019	12.5	101	0.00
37	2-Methylnaphthalene	40.000	34.429	13.9	101	0.00
38	1-Methylnaphthalene	40.000	34.689	13.3	101	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	101	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	36.051	9.9	103	0.00
41 P	Hexachlorocyclopentadiene	40.000	36.655	8.4	100	0.00
42 S	2,4,6-Tribromophenol	80.000	75.593	5.5	107	0.00
43 C	2,4,6-Trichlorophenol	40.000	36.571	8.6	103	0.00
44	2,4,5-Trichlorophenol	40.000	36.346	9.1	103	0.00
45 S	2-Fluorobiphenyl	80.000	69.788	12.8	103	0.00
46	1,1'-Biphenyl	40.000	34.676	13.3	100	0.00
47	2-Chloronaphthalene	40.000	35.038	12.4	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF011822\
 Data File : BF126602.D
 Acq On : 18 Jan 2022 13:37
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF011822

Quant Time: Jan 19 07:05:54 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 19 07:04:14 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.653	-4.1	107	0.00
49	Acenaphthylene	40.000	34.853	12.9	101	0.00
50	Dimethylphthalate	40.000	34.834	12.9	100	0.00
51	2,6-Dinitrotoluene	40.000	40.911	-2.3	107	0.00
52 C	Acenaphthene	40.000	35.448	11.4	102	0.00
53	3-Nitroaniline	40.000	40.074	-0.2	106	0.00
54 P	2,4-Dinitrophenol	40.000	43.470	-8.7	120	0.00
55	Dibenzofuran	40.000	35.147	12.1	103	0.00
56 P	4-Nitrophenol	40.000	40.590	-1.5	106	0.00
57	2,4-Dinitrotoluene	40.000	43.401	-8.5	111	0.00
58	Fluorene	40.000	34.871	12.8	102	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	37.167	7.1	103	0.00
60	Diethylphthalate	40.000	35.143	12.1	101	0.00
61	4-Chlorophenyl-phenylether	40.000	35.026	12.4	103	0.00
62	4-Nitroaniline	40.000	39.710	0.7	105	0.00
63	Azobenzene	40.000	34.635	13.4	100	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	101	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.906	-2.3	117	0.00
66 c	n-Nitrosodiphenylamine	40.000	34.882	12.8	101	0.00
67	4-Bromophenyl-phenylether	40.000	35.882	10.3	104	0.00
68	Hexachlorobenzene	40.000	35.979	10.1	104	0.00
69	Atrazine	40.000	34.973	12.6	101	0.00
70 C	Pentachlorophenol	40.000	38.041	4.9	105	0.00
71	Phenanthrene	40.000	34.435	13.9	101	0.00
72	Anthracene	40.000	34.305	14.2	101	0.00
73	Carbazole	40.000	34.306	14.2	101	0.00
74	Di-n-butylphthalate	40.000	34.751	13.1	100	0.00
75 C	Fluoranthene	40.000	34.554	13.6	100	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	99	0.00
77	Benzidine	40.000	30.053	24.9	90	0.00
78	Pyrene	40.000	35.886	10.3	100	0.00
79 S	Terphenyl-d14	80.000	72.076	9.9	102	0.00
80	Butylbenzylphthalate	40.000	37.184	7.0	101	0.00
81	Benzo(a)anthracene	40.000	35.816	10.5	101	0.00
82	3,3'-Dichlorobenzidine	40.000	34.899	12.8	98	0.00
83	Chrysene	40.000	35.509	11.2	102	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	36.677	8.3	101	0.00
85 c	Di-n-octyl phthalate	40.000	35.249	11.9	101	0.00
86 I	Perylene-d12	20.000	20.000	0.0	103	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	38.190	4.5	106	0.00
88	Benzo(b)fluoranthene	40.000	35.517	11.2	102	0.00
89	Benzo(k)fluoranthene	40.000	36.113	9.7	100	0.00
90 C	Benzo(a)pyrene	40.000	35.426	11.4	100	0.00
91	Dibenzo(a,h)anthracene	40.000	37.934	5.2	104	0.00
92	Benzo(g,h,i)perylene	40.000	38.619	3.5	106	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF011822\
Data File : BF126602.D
Acq On : 18 Jan 2022 13:37
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF011822

Quant Time: Jan 19 07:05:54 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 19 07:04:14 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0