


Data Path : Z:\HPCHEM1\BNA F\DATA\BF011916\  
 Data File : BF084543.D  
 Acq On : 19 Jan 2016 9:02  
 Operator : SJ/UM  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF123115.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Thu Jan 14 14:15:52 2016



AutoFind: Scans 830, 831, 832; Background Corrected with Scan 827

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 42.1      | 48398   | PASS             |
| 68          | 69           | 0.00         | 2            | 1.9       | 816     | PASS             |
| 69          | 198          | 0.00         | 100          | 36.6      | 42057   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.3       | 147     | PASS             |
| 127         | 198          | 10           | 80           | 51.8      | 59586   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.5       | 536     | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 114970  | PASS             |
| 199         | 198          | 5            | 9            | 6.6       | 7638    | PASS             |
| 275         | 198          | 10           | 60           | 22.9      | 26279   | PASS             |
| 365         | 198          | 1            | 100          | 2.4       | 2816    | PASS             |
| 441         | 442          | 0.01         | 24           | 16.9      | 15160   | PASS             |
| 442         | 198          | 50           | 100          | 78.2      | 89928   | PASS             |
| 443         | 442          | 15           | 24           | 19.5      | 17547   | PASS             |