

Data Path : Z:\HPCHEM1\BNA F\DATA\BF011918\
 Data File : BF102239.D
 Acq On : 20 Jan 2018 1:32
 Operator : SJ/JU
 Sample : J1189-03
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 3D-8

Quant Time: Jan 20 02:44:12 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF010918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 19 13:31:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.72	152	148654	20.00	ng	0.00
21) Naphthalene-d8	8.00	136	611287	20.00	ng	-0.01
38) Acenaphthene-d10	9.76	164	269747	20.00	ng	0.00
63) Phenanthrene-d10	11.24	188	428897	20.00	ng	0.00
75) Chrysene-d12	13.87	240	275177	20.00	ng	0.00
86) Perylene-d12	15.29	264	284061	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.33	112	981928	93.27	ng	0.00
7) Phenol-d6	6.36	99	1176032	93.63	ng	0.00
23) Nitrobenzene-d5	7.29	82	730651	70.24	ng	-0.01
41) 2,4,6-Tribromophenol	10.55	330	243034	103.24	ng	0.00
44) 2-Fluorobiphenyl	9.08	172	1153132	66.78	ng	0.00
78) Terphenyl-d14	12.82	244	837668	63.20	ng	0.00
Target Compounds						
10) Phenol	6.37	94	43343	3.054	ng	94
37) 2-Methylnaphthalene	8.72	142	89642	4.458	ng	97
49) Dimethylphthalate	9.47	163	261942	12.105	ng	98
70) Phenanthrene	11.26	178	84855	3.387	ng	98
83) Bis(2-ethylhexyl)phthalate	13.86	149	36879	2.880	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF011918\
Data File : BF102239.D
Acq On : 20 Jan 2018 1:32
Operator : SJ/JU
Sample : J1189-03
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
3D-8

Quant Time: Jan 20 02:44:12 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF010918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 19 13:31:35 2018
Response via : Initial Calibration

