


Data Path : Z:\HPCHEM1\BNA F\Data\BF011918\
 Data File : BF102225.D
 Acq On : 19 Jan 2018 18:48
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF010918.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Jan 19 13:31:35 2018



AutoFind: Scans 1598, 1599, 1600; Background Corrected with Scan 1594

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.8	39000	PASS
68	69	0.00	2	2.0	672	PASS
69	198	0.00	100	37.9	33728	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	51.9	46200	PASS
197	198	0.00	2	0.7	593	PASS
198	198	100	100	100.0	88981	PASS
199	198	5	9	6.5	5819	PASS
275	198	10	60	22.6	20146	PASS
365	198	1	100	2.8	2520	PASS
441	443	0.01	100	81.3	12001	PASS
442	198	50	100	86.8	77216	PASS
443	442	15	24	19.1	14755	PASS