

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF011924\
 Data File : BF137115.D
 Acq On : 19 Jan 2024 16:54
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF011924

Quant Time: Jan 20 00:58:58 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jan 20 00:58:12 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	97	0.00
2	1,4-Dioxane	40.000	40.567	-1.4	97	0.00
3	Pyridine	40.000	41.573	-3.9	99	0.00
4	n-Nitrosodimethylamine	40.000	40.914	-2.3	97	0.00
5 S	2-Fluorophenol	80.000	80.202	-0.3	96	0.00
6	Aniline	40.000	42.686	-6.7	96	0.00
7 S	Phenol-d6	80.000	80.601	-0.8	97	0.00
8	2-Chlorophenol	40.000	40.732	-1.8	97	0.00
9	Benzaldehyde	40.000	40.489	-1.2	104	0.00
10 C	Phenol	40.000	41.374	-3.4	98	0.00
11	bis(2-Chloroethyl)ether	40.000	39.111	2.2	96	0.00
12	1,3-Dichlorobenzene	40.000	40.308	-0.8	96	0.00
13 C	1,4-Dichlorobenzene	40.000	40.319	-0.8	97	0.00
14	1,2-Dichlorobenzene	40.000	40.386	-1.0	97	0.00
15	Benzyl Alcohol	40.000	40.929	-2.3	96	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.921	0.2	95	0.00
17	2-Methylphenol	40.000	40.725	-1.8	98	0.00
18	Hexachloroethane	40.000	41.091	-2.7	97	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.411	-1.0	98	0.00
20	3+4-Methylphenols	40.000	40.610	-1.5	97	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	97	0.00
22	Acetophenone	40.000	39.874	0.3	98	0.00
23 S	Nitrobenzene-d5	80.000	87.478	-9.3	101	0.00
24	Nitrobenzene	40.000	42.579	-6.4	98	0.00
25	Isophorone	40.000	39.993	0.0	97	0.00
26 C	2-Nitrophenol	40.000	41.971	-4.9	103	0.00
27	2,4-Dimethylphenol	40.000	40.125	-0.3	97	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.738	0.7	95	0.00
29 C	2,4-Dichlorophenol	40.000	41.300	-3.2	99	0.00
30	1,2,4-Trichlorobenzene	40.000	40.302	-0.8	96	0.00
31	Naphthalene	40.000	40.156	-0.4	97	0.00
32	Benzoic acid	40.000	43.794	-9.5	102	0.00
33	4-Chloroaniline	40.000	38.956	2.6	97	0.00
34 C	Hexachlorobutadiene	40.000	40.602	-1.5	97	0.00
35	Caprolactam	40.000	40.604	-1.5	97	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.587	-1.5	98	0.00
37	2-Methylnaphthalene	40.000	40.214	-0.5	97	0.00
38	1-Methylnaphthalene	40.000	40.439	-1.1	99	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	98	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.137	-0.3	98	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.666	-1.7	92	0.00
42 S	2,4,6-Tribromophenol	80.000	82.849	-3.6	101	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.664	-1.7	97	0.00
44	2,4,5-Trichlorophenol	40.000	41.144	-2.9	98	0.00
45 S	2-Fluorobiphenyl	80.000	79.288	0.9	98	0.00
46	1,1'-Biphenyl	40.000	40.240	-0.6	96	0.00
47	2-Chloronaphthalene	40.000	40.537	-1.3	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF011924\
 Data File : BF137115.D
 Acq On : 19 Jan 2024 16:54
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF011924

Quant Time: Jan 20 00:58:58 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jan 20 00:58:12 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	46.523	-16.3	102	0.00
49	Acenaphthylene	40.000	40.595	-1.5	97	0.00
50	Dimethylphthalate	40.000	40.176	-0.4	98	0.00
51	2,6-Dinitrotoluene	40.000	45.612	-14.0	102	0.00
52 C	Acenaphthene	40.000	40.481	-1.2	98	0.00
53	3-Nitroaniline	40.000	43.765	-9.4	99	0.00
54 P	2,4-Dinitrophenol	40.000	40.756	-1.9	109	0.00
55	Dibenzofuran	40.000	40.179	-0.4	97	0.00
56 P	4-Nitrophenol	40.000	44.411	-11.0	99	0.00
57	2,4-Dinitrotoluene	40.000	47.629	-19.1	103	0.00
58	Fluorene	40.000	39.856	0.4	99	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.140	-0.4	97	0.00
60	Diethylphthalate	40.000	40.597	-1.5	98	0.00
61	4-Chlorophenyl-phenylether	40.000	39.406	1.5	99	0.00
62	4-Nitroaniline	40.000	44.897	-12.2	101	0.00
63	Azobenzene	40.000	40.171	-0.4	97	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	98	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.601	-1.5	105	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.959	0.1	97	0.00
67	4-Bromophenyl-phenylether	40.000	40.695	-1.7	98	0.00
68	Hexachlorobenzene	40.000	39.422	1.4	97	0.00
69	Atrazine	40.000	42.488	-6.2	97	0.00
70 C	Pentachlorophenol	40.000	42.001	-5.0	97	0.00
71	Phenanthrene	40.000	40.047	-0.1	98	0.00
72	Anthracene	40.000	40.250	-0.6	97	0.00
73	Carbazole	40.000	39.917	0.2	97	0.00
74	Di-n-butylphthalate	40.000	40.949	-2.4	97	0.00
75 C	Fluoranthene	40.000	39.674	0.8	97	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	98	0.00
77	Benzydine	40.000	53.086	-32.7#	90	0.00
78	Pyrene	40.000	41.559	-3.9	96	0.00
79 S	Terphenyl-d14	80.000	81.816	-2.3	98	0.00
80	Butylbenzylphthalate	40.000	44.237	-10.6	101	0.00
81	Benzo(a)anthracene	40.000	41.236	-3.1	100	0.00
82	3,3'-Dichlorobenzidine	40.000	40.381	-1.0	97	0.00
83	Chrysene	40.000	40.617	-1.5	98	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	44.616	-11.5	101	0.00
85 c	Di-n-octyl phthalate	40.000	43.254	-8.1	103	0.00
86 I	Perylene-d12	20.000	20.000	0.0	103	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	43.776	-9.4	100	0.00
88	Benzo(b)fluoranthene	40.000	37.802	5.5	96	0.00
89	Benzo(k)fluoranthene	40.000	40.804	-2.0	107	0.00
90 C	Benzo(a)pyrene	40.000	40.551	-1.4	102	0.00
91	Dibenzo(a,h)anthracene	40.000	43.616	-9.0	100	0.00
92	Benzo(g,h,i)perylene	40.000	43.979	-9.9	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF011924\
Data File : BF137115.D
Acq On : 19 Jan 2024 16:54
Operator : CG\JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF011924

Quant Time: Jan 20 00:58:58 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF011924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jan 20 00:58:12 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0