


Data Path : Z:\HPCHEM1\BNA F\DATA\BF012015\
 Data File : BF076979.D
 Acq On : 21 Jan 2015 00:47
 Operator : IZ / TP
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF011415.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Jan 21 11:38:13 2015



AutoFind: Scans 1507, 1508, 1509; Background Corrected with Scan 1504

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.9	5546	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	39.3	4955	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	55.8	7047	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	12620	PASS
199	198	5	9	7.3	915	PASS
275	198	10	60	24.1	3039	PASS
365	198	1	100	1.3	170	PASS
441	442	0.01	24	15.6	1731	PASS
442	198	50	100	88.0	11107	PASS
443	442	15	24	17.6	1955	PASS