

Data Path : Z:\HPCHEM1\BNA F\DATA\BF012017\
 Data File : BF092383.D
 Acq On : 20 Jan 2017 20:10
 Operator : UM/SJ
 Sample : I1265-01 5X
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 B-1

Manual Integrations
 APPROVED

mohammad
 1/23/2017 2:45:03 PM

Quant Time: Jan 23 11:59:47 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF122116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 20 22:22:14 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.52	152	185979	20.00	ng	0.00
21) Naphthalene-d8	7.81	136	711054	20.00	ng	0.00
38) Acenaphthene-d10	9.56	164	323547	20.00	ng	0.00
63) Phenanthrene-d10	11.03	188	444310	20.00	ng	0.00
75) Chrysene-d12	13.66	240	329597	20.00	ng	0.00
86) Perylene-d12	15.02	264	240124	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.12	112	34831	3.13	ng	0.03
7) Phenol-d6	6.21	99	47090	3.46	ng	0.02
23) Nitrobenzene-d5	7.10	82	31920	2.50	ng	0.00
41) 2,4,6-Tribromophenol	10.36	330	8427	3.15	ng	0.01
44) 2-Fluorobiphenyl	0.00	172	0d	0.00	ng	
78) Terphenyl-d14	12.62	244	52380	3.85	ng	0.00
Target Compounds						Qvalue
77) Pyrene	12.47	202	128666	5.32	ng	97
80) Benzo(a)anthracene	13.65	228	75374	4.05	ng	97
82) Chrysene	13.70	228	56622m	3.03	ng	
87) Benzo(b)fluoranthene	14.67	252	84500m	5.65	ng	
89) Benzo(a)pyrene	14.97	252	51411	3.87	ng	# 90
91) Benzo(g,h,i)perylene	16.60	276	31718	2.80	ng	# 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF012017\
Data File : BF092383.D
Acq On : 20 Jan 2017 20:10
Operator : UM/SJ
Sample : I1265-01 5X
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
B-1

Manual Integrations
APPROVED

mohammad
1/23/2017 2:45:03 PM

Quant Time: Jan 23 11:59:47 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF122116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 20 22:22:14 2017
Response via : Initial Calibration

