

Data Path : Z:\HPCHEM1\BNA F\DATA\BF012017\
 Data File : BF092389.D
 Acq On : 20 Jan 2017 23:02
 Operator : UM/SJ
 Sample : I1329-07
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 S2-0-3FT-COMP

Manual Integrations
 APPROVED

mohammad
 1/23/2017 2:45:20 PM

Quant Time: Jan 21 00:18:18 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF122116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 20 22:22:14 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.52	152	139492	20.00	ng	0.00
21) Naphthalene-d8	7.81	136	702905	20.00	ng	0.00
38) Acenaphthene-d10	9.56	164	324300	20.00	ng	0.00
63) Phenanthrene-d10	11.03	188	439849	20.00	ng	0.00
75) Chrysene-d12	13.66	240	318418	20.00	ng	0.00
86) Perylene-d12	15.02	264	240714	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.11	112	236875	28.35	ng	0.02
7) Phenol-d6	6.19	99	899256	88.17	ng	0.00
23) Nitrobenzene-d5	7.10	82	833260	65.92	ng	0.00
41) 2,4,6-Tribromophenol	10.35	330	26329	9.81	ng	0.00
44) 2-Fluorobiphenyl	8.90	172	1444478	92.38	ng	0.00
78) Terphenyl-d14	12.62	244	901017	68.63	ng	0.00
Target Compounds						Qvalue
10) Phenol	6.21	94	51121	4.40	ng	87
30) 1,2,4-Trichlorobenzene	7.76	180	33426	3.27	ng	99
31) Naphthalene	7.83	128	135360	4.21	ng	99
37) 2-Methylnaphthalene	8.53	142	79615	2.72	ng	97
45) 1,1'-Biphenyl	8.99	154	53199	2.40	ng	94
49) Dimethylphthalate	9.30	163	355156	17.12	ng	100
70) Phenanthrene	11.06	178	174801	7.33	ng	98
74) Fluoranthene	12.25	202	191179	5.81	ng	94
77) Pyrene	12.47	202	180392	7.72	ng	97
80) Benzo(a)anthracene	13.66	228	88445	4.92	ng	95
82) Chrysene	13.70	228	76292m	4.23	ng	
83) Bis(2-ethylhexyl)phthalate	13.67	149	34002	2.72	ng	# 89
87) Benzo(b)fluoranthene	14.67	252	87199m	5.82	ng	
89) Benzo(a)pyrene	14.97	252	49013	3.68	ng	# 74
91) Benzo(a,h,i)perylene	16.60	276	27773	2.44	ng	# 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF012017\
Data File : BF092389.D
Acq On : 20 Jan 2017 23:02
Operator : UM/SJ
Sample : I1329-07
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
S2-0-3FT-COMP

Manual Integrations
APPROVED

mohammad
1/23/2017 2:45:20 PM

Quant Time: Jan 21 00:18:18 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF122116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 20 22:22:14 2017
Response via : Initial Calibration

