


Data Path : Z:\HPCHEM1\BNA F\DATA\BF012017\
 Data File : BF092365.D
 Acq On : 20 Jan 2017 10:51
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF122116.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Jan 06 15:43:51 2017



AutoFind: Scans 833, 834, 835; Background Corrected with Scan 829

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.3	35720	PASS
68	69	0.00	2	1.5	544	PASS
69	198	0.00	100	35.6	37061	PASS
70	69	0.00	2	0.7	257	PASS
127	198	10	80	47.2	49162	PASS
197	198	0.00	2	0.6	605	PASS
198	198	100	100	100.0	104134	PASS
199	198	5	9	6.7	6936	PASS
275	198	10	60	24.0	24967	PASS
365	198	1	100	2.5	2616	PASS
441	442	0.01	24	15.6	14483	PASS
442	198	50	100	88.9	92608	PASS
443	442	15	24	19.3	17870	PASS