


Data Path : Z:\HPCHEM1\BNA\_F\Data\BF012018\  
 Data File : BF102278.D  
 Acq On : 20 Jan 2018 22:36  
 Operator : SJ/JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF010918.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Fri Jan 19 13:31:35 2018



AutoFind: Scans 1598, 1599, 1600; Background Corrected with Scan 1593

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 39.6      | 25519   | PASS             |
| 68          | 69           | 0.00         | 2            | 1.3       | 307     | PASS             |
| 69          | 198          | 0.00         | 100          | 35.3      | 22762   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 127         | 198          | 10           | 80           | 49.2      | 31720   | PASS             |
| 197         | 198          | 0.00         | 2            | 1.1       | 685     | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 64469   | PASS             |
| 199         | 198          | 5            | 9            | 6.7       | 4293    | PASS             |
| 275         | 198          | 10           | 60           | 22.4      | 14473   | PASS             |
| 365         | 198          | 1            | 100          | 2.6       | 1681    | PASS             |
| 441         | 443          | 0.01         | 100          | 79.9      | 8501    | PASS             |
| 442         | 198          | 50           | 100          | 86.4      | 55698   | PASS             |
| 443         | 442          | 15           | 24           | 19.1      | 10635   | PASS             |